OCT 16, 2019

🌐 OD and GFP Plate Reader Assay (72 h Measurement)

Alba Balletbó¹

¹Wageningen University

iGEM Wageningen 2019

Alba Balletbó

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.8aihsce

Protocol Citation: Alba Balletbó 2019. OD and GFP Plate Reader Assay (72 h Measurement). **protocols.io** <https://dx.doi.org/10.17504/protocols.io.8aihsce>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Oct 16, 2019

Last Modified: Oct 16, 2019

PROTOCOL integer ID:
 28714

ABSTRACT

Fluorescence measurements of Optical Density (OD) and Green fluorescent protein (GFP) in *Escherichia coli*.

MATERIALS

MATERIALS

Microplate Reader Synergy Mx **Contributed by users**

Nunc™ MicroWell™ 96-Well Microplates, F 96 well plate, cell culture, clear, with lid, Sterile **Thermo Fisher Catalog #167008**

Media Preparation

- 1 Prepare the following media and autoclave them according to standard procedures.

M9TG Media (Without carbon source):

Reagent	Amount to add for 50 mL
M9 Salts	1X
Tryptone	0.5 g

M9TG Media (0.5% Glycerol):

Reagent	Amount to add for 50 mL
M9 Salts	1X
Tryptone	0.5 g
Glycerol	0.25 g

M9TG Media (2% Glycerol):

Reagent	Amount to add for 50 mL
M9 Salts	1X
Tryptone	0.5 g
Glycerol	1 g

M9TG Media (0.5% Succinate and 1mM Leucine):

Reagent	Amount to add for 50 mL
M9 Salts	1X
Tryptone	0.5 g
Succinate	0.25 g
Leucine	6.56 mg

Bacteria and Plate Reader Preparation

- 2 Grow desired bacteria overnight in a 50 ml falcon tub, containing 10 mL M9TG media and grown cultures @ 37 °C, 250 RPM overnight.

Note

Luria-Bertani media is also possible to use, but cells must be washed with PBS buffer prior to plate reader experiment because of the high amount of fluorescence from the yeast extract:

After overnight incubation, spin cells down at 4700 x G, for 5 min, discard the supernatant and resuspend in 10 ml PBS.

Repeat prior steps and resuspension again in PBS.

- 3 Load 190 µL of the corresponding buffer in the wells of the 96-well plate and add 10 µL of overnight culture.

Note

The amount of buffer and overnight culture will depend on the OD after overnight bacterial cultivation. The amounts present in this protocol were determined by an OD ~0.3.

Note

When cells have been grown overnight in LB media, washing steps are required before continuing! (*See step 3*).

Plate Reader Protocol

- 4 Set the following variables for the BioTek Synergy MX Microplate Reader:

- Temperature: Setpoint 37°C

Preheat before moving to next step. Start Kinetic [Run 72:00:00, Interval 0:04:20]

- Shake: Medium for 0:05 (MM:SS)
- **Read OD600:**
 - Absorbance Endpoint of full plate
 - Wavelengths: 600
 - Read Speed: Normal
- **Read GFP:**
 - Fluorescence Endpoint of full plate
 - Filter Set:
 - Excitation: 488/9,0
 - Emission: 510/9,0
 - Optics: Bottom
 - Gain: 100
 - Read Speed: Normal

5 Start plate reader protocol.

Data Analysis

6 Retrieve the data from the computer, correct for the measurements for the OD600 and subtract the autofluorescence from the control sample of all the other samples.