NOV 23, 2018

Optical density measurement for Bacterial solution (Worm Food)

Priota Islam¹

¹Imperial College London

Behavioural Genomics

Priota Islam
Imperial College London

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.vude6s6

Protocol Citation: Priota Islam 2018. Optical density measurement for Bacterial solution (Worm Food).
protocols.io
<https://dx.doi.org/10.17504/protocols.io.vude6s6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Nov 23, 2018

Last Modified: Nov 23, 2018

PROTOCOL integer ID:
18021

ABSTRACT

Before the OP50 solution is aliquoted into small falcon tubes, it is always better to check the OD by using a spectrophotometer, for documentation purposes.

1 Pre-measurement

1. Clean the disposable cuvettes with ethanol and lint free paper
2. Aliquot small amount of LB broth into a falcon tube (For Blank)

3. Fill about 1ml of the LB broth into a clean cuvette
4. Mix/Vortex the tube containing the stock bacteria to achieve a homogenized solution
5. Take three aliquots from the same tube in three different cuvettes

2 **Actual measurement**

1. Switch on the spectrophotometer
2. Press the bottom left button to select the OD program
3. Increase the wavelength to 600nm
4. Put the blank cuvette first and press the button showing blank
5. Put the sample cuvette 1 and press the button showing sample
6. Take a note of the OD reading
7. Place the second sample and repeat
8. After taking readings for all three samples take the average
9. Record the OD as OD₆₀₀ = 1.35

3 **Post measurement**

1. Discard the cuvettes
2. Close the lid to the spectrophotometer
3. Switch it OFF