OCT 29, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.wxqffmw

Protocol Citation: Liu Liu 2019. Painting Neuropixels probes and other silicon probes for electrophysiological recordings. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.wxqffmw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Jan 09, 2019

Last Modified: Oct 29, 2019

PROTOCOL integer ID:
 19152

Keywords: Neuropixels probes, silicon probes, electrode track, dye

Painting Neuropixels probes and other silicon probes for electrophysiological recordings

Liu Liu¹

¹Janelia Research Campus, HHMI

Liu Liu
 Janelia Research Campus

ABSTRACT

This protocol is for consistently applying fluorescent dyes to Neuropixels probes and other types of silicon probes, which allows imaging and reconstruction of the electrode tracks in the brain in post hoc histology.

MATERIALS

MATERIALS

Microscope slides **Contributed by users**

Coverslip **Contributed by users**

CM-Dil **Thermo Fisher Scientific Catalog #C7000**

The CM-Dil should be stored in the freezer at -20°C. Before using, dissolve a 50 µg vial in 50 µL of pure ethanol. Tightly seal the vial after using to prevent the ethanol from evaporating. The remaining CM-Dil can be stored at room temperature until use. Protect from light.

Secure Neuropixels probes (shown) or other silicon probes onto a micromanipulator. The side

- 1 with eletrode sites should face up.
- 2 Secure a microscope slide onto a holder, and a clean coverslip at the end of the microscope slide. Place the coverslip attached to the microscope slide near to the Neuropixels probe (2-3 mm).
- 3 Using a micropipette, place 1 μL of CM-Dil solution close to the edge of the coverslip. To ensure a consistent labeling of the probe, make sure the pipette does not produce an air bubble in the drop of dye. The dye concentration should be 1 $\mu\text{g}/\mu\text{L}$ CM-Dil in ethanol. The drop of dye will evaporate in ~ 30 s at room temperature, so it is important to complete step 4 quickly.
- 4 This step should be done under a microscope. Using the micromanipulator, dip the tip of the Neuropixels probe into the dye. Be careful not to get dye onto the electrode sites. For Neuropixels probes, the tip extends only about 150 μm from the first electrode site (see photo).

Neuropixels probe shank showing the electrodes sites and the tip.

The electrodes sites should

be visible under the microscope. The experimenter should be familiar with the setup to comfortably complete this step in considerably less than 30 s.

- 5 Wait until ethanol evaporates and dye dries onto the tip of the probe, around 30 seconds for ethanol-based CM-Dil.
- 6 Using the micromanipulator, slowly retract the probe away from the coverslip and dismount. One

should be able to see dried pink dye on the tip of the probe. If not, repeat painting the probe.