VERSION 1

OCT 27, 2017

Parechovirus real-time RT-PCR V.1

Ian M
Mackay¹

¹The University of Queensland

Ian M Mackay
The University of Queensland

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.kgnctve

Protocol Citation: Ian M Mackay 2017. Parechovirus real-time RT-PCR.

protocols.io

<https://dx.doi.org/10.17504/protocols.io.kgnctve>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

Created: Oct 27, 2017

Last Modified: Mar 27, 2018

PROTOCOL integer ID:
8430

Oligonucleotide sequences...

1 *AN345_panHPeV/LV (sense primer)*

GTAACASWWGCCTCTGGGSCCAAAAG

AN344_panHPeV/LV (antisense probe)

GGCCCCWGRTCAGATCCAYAGT

AN257_HPeV/LV

FAM-CCTRYGGGTACCTYCWGGGCATCCTTC-BHQ1

Reaction setup...

- 2 Below is the reaction setup for a single RT-PCR reaction.

Ideally, this work is conducted in a laboratory separate to any space used to perform PCR, molecular cloning or the analysis of high concentration DNA.

This volume has been used in 0.1-0.2ml tubes or various other connected tube configurations such as 100-place rings.

Multiply this according to the number of reactions you will need, remembering to include a positive control and at least two non-template controls (NTCs)

You may also need to allow some extra volume, depending on the method used to pipette mix into tubes for the run. For example, some robot-loaded tubes can require two reaction "dead volumes".

<i>Reagent (stock concentration)</i>	<i>Vol (μL) / reaction</i>	<i>Final concentration</i>
Nuclease free water	4.47	N/A
AN345_panHPeV/LV (200pmol/ul)	0.03	300nM
AN344_panHPeV/LV (200pmol/ul)	0.03	300nM
AN257_HPeV/LV FAM-BHQ1 (100pmol/ul)	0.03	150nM
2X Reaction Mix ¹	10	1X
Rox Reference Dye 25mM ¹	0.04	50nM
SuperScript® III/Platinum® Taq Mix ¹	0.4	1X
Template extract RNA	5	N/A
Final volume	20μl	

¹SuperScript® III Platinum® One-Step qRT-PCR Kit, Cat No. 11732088

Amplification...

This assay has been optimized and validated for use with a Rotor-Gene 6000 or Rotor-Gene Q

3 thermal cycler.

The cycling conditions are as follows:

<i>R T - P C R</i>
5 0 ° C
5 m in
9 5 ° C
2 m in
9 5 ° C
3 s
4 0 X 6 0 ° C
3 0 s *

*Florescence acquisition step

Result calling...

- 4 The definition used for a satisfactory positive result from a real-time fluorogenic PCR should include each of the following:
 1. A **sigmoidal curve** – the trace travels horizontally, curves upward, continues in an exponential

- rise and followed by a curve towards a horizontal plateau phase
2. A **suitable level of fluorescence** intensity as measured in comparison to controls using (y-axis)
3. A defined threshold (C_T) value which the fluorescent curve has clearly exceeded (Fig.1 arrow) and which sits early in the log-linear phase
4. A flat or non-sigmoidal curve or a curve that crosses the threshold with a C_T value >40 cycles is considered a negative result

Figure 1. Examples of satisfactory sigmoidal amplification curve shape when considering an assay's fluorescent signal output. The crossing point or threshold cycle (C_T) is indicated (yellow arrow); it is the value at which fluorescence levels surpass a predefined (usually set during validation, or arbitrary) threshold level as shown in this normalized linear scale depiction. LP-log-linear phase of signal generated during the exponential part of the PCR amplification; TP-a slowing of the amplification and accompanying fluorescence signal marks the transition phase; PP-the plateau phase is reached when there is little or no increase in fluorescent signal despite continued cycling.

References...

- 5 The oligonucleotides used in this assay have been previously published.
 1. Nix WA, Maher K, Johansson ES, Niklasson B, Lindberg AM, Pallansch MA, Oberste MS. Detection of all known parechoviruses by real-time PCR. J Clin Microbiol.46(8):2519-24.