SEP 20, 2019

PCR

宏亮 董¹

¹Northeast Forest University

2019 iGEM NEFU_China

Tech. support email: shengyiyanwork@gmail.com

宏亮 董

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.7hyhj7w

Protocol Citation: 宏亮 董
2019. PCR. [protocols.io](https://dx.doi.org/10.17504/protocols.io.7hyhj7w)
<https://dx.doi.org/10.17504/protocols.io.7hyhj7w>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Sep 20, 2019

Last Modified: Sep 20, 2019

PROTOCOL integer ID:
27928

Keywords: PCR

MATERIALS

2×high Taq Master Mix

ddH₂O

Template

F/R Primer

Bio-rad S1000TM Thermo Cycler.

SAFETY WARNINGS

Please wear gloves for the experiment, don't try to touch the lid after PCR program initiation.

BEFORE START INSTRUCTIONS

Synthesize primers in advance.

Before the start

- 1 Set up a small box with ice, put DNA and 2×high Taq Master Mix into it before going into the Bio-

rad S1000TM thermocycler. Set up a small box with ice, put DNA and 2×high Taq Master Mix into it before going into the Bio-rad S1000TM Thermo Cycler.

2 Add the following reagent to a PCR tube. (50 µl).

2×high Taq Master Mix (Enzyme)	25 µl
Template	2 µl
Forward Primer (10 µM)	2 µl
Reverse Primer (10 µM)	2 µl
ddH ₂ O	19 µl

3 Program the thermocycler as follows:

Temperature	Time
98°C	5 min
98°C	15 s
T _m -4°C	15 s
72°C	(5-10s) × (1 kb)
72°C	5 min
4°C	∞