OCT 24, 2020

PCR

Zhujun Wei¹

¹Northeast Forest University

2020 iGEM NEFU China

Zhujun Wei

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bnwqmfdw

Protocol Citation: Zhujun Wei 2020. PCR. [protocols.io](https://dx.doi.org/10.17504/protocols.io.bnwqmfdw)
<https://dx.doi.org/10.17504/protocols.io.bnwqmfdw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Oct 24, 2020

Last Modified: Oct 24, 2020

PROTOCOL integer ID:
 43696

MATERIALS

2×high Taq Master Mix
 ddH₂O
 Template
 F/R Primer
 Bio-rad S1000TM Thermo Cycler.

SAFETY WARNINGS

Please wear gloves for the experiment, don't try to touch the lid after PCR program initiation.

BEFORE START INSTRUCTIONS

Set up a small box with ice, put DNA and 2×high Taq Master Mix into it before going into the Bio-rad S1000TM thermocycler. Set up a small box with ice, put DNA and 2×high Taq Master Mix into it before going into the Bio-rad S1000TM Thermo Cycler.

1 Program the thermocycler as follows:

Temperature	Time
98°C	5 min

98°C	15 s
Tm-4°C	15 s
72°C	(5-10s)× (1 kb)
72°C	5 min
4°C	∞

2 Add the following reagent to a PCR tube.(50 µl).

2×high Taq Master Mix (Enzyme)	25 µl
Template	2 µl
Forward Primer (10 µM)	2 µl
Reverse Primer (10 µM)	2 µl
ddH2O	19 µl