

Nov 26, 2020

Performance Study of Wireless Fecobionics Device in Canine

DOI

dx.doi.org/10.17504/protocols.io.bpzfmp3n

Yanmin Wang¹, Hans Gregersen¹

¹California Medical Innovations Institute



Yanmin Wang

OPEN  ACCESS



DOI: dx.doi.org/10.17504/protocols.io.bpzfmp3n

Protocol Citation: Yanmin Wang, Hans Gregersen 2020. Performance Study of Wireless Fecobionics Device in Canine. [protocols.io](https://dx.doi.org/10.17504/protocols.io.bpzfmp3n) <https://dx.doi.org/10.17504/protocols.io.bpzfmp3n>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: November 24, 2020

Last Modified: November 26, 2020

Protocol Integer ID: 44807

Keywords: Colon, bionics, cannula, colonoscopy, fluoroscopy,

Abstract

We developed a novel wireless device (Fecobionics) for mapping colonic and anorectal neuromuscular function. The hypothesis of this protocol is that the Fecobionics device can test pressures, orientation, bending, shape, and cross-sectional area changes in colon and rectum. To validate our hypothesis, we inserted the device to the proximal colon, either through a cannula or colonoscopy, followed by data transmission and recording.

Cannula test

- 1 After a laparotomy, a cannula was implanted into the proximal of the colon (4-5 cm to cecum)
- 2 The external end of the cannula was screwed by a cap, which is able to open for device insertion
- 3 After 10-14 days recovery, the Fecobionics was inserted through the cannula for testing
- 4 X-ray fluoroscopy was performed frequently to locate the device. Data was recorded for analysis.

Colonoscopy

- 5 After 2-day laxative and an enema, the animal was anesthetize for colonoscopy
- 6 The Fecobionics was delivered to proximal colon (close to cecum) by an endoscopic rat-tooth forcep
- 7 After the delivery, X-ray fluoroscopy was performed frequently to locate the device. Data was recorded for analysis.