OCT 29, 2019

OPEN ACCESS

DOI:
[dx.doi.org/10.17504/protocols.io.6zxhf7n](https://doi.org/10.17504/protocols.io.6zxhf7n)

External link:
<https://doi.org/10.1371/journal.pone.0224276>

Protocol Citation: John Davis Coakley 2019. Phenotypic analysis of circulating leukocytes. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.6zxhf7n>

MANUSCRIPT CITATION:
 Coakley JD, Breen EP, Moreno-Olivera A, Al-Harbi AI, Melo AM, O'Connell B, McManus R, Doherty DG, Ryan T (2019) Dysregulated T helper type 1 (Th1) and Th17 responses in elderly hospitalised patients with infection and sepsis. PLoS ONE 14(10): e0224276. doi: [10.1371/journal.pone.0224276](https://doi.org/10.1371/journal.pone.0224276)

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

Phenotypic analysis of circulating leukocytes

PLOS One

John Davis Coakley¹

¹Trinity College Dublin

John Davis Coakley

ABSTRACT

Antibody panels for flow cytometry were prepared with specific antibodies for identification of lymphocyte populations, and antibodies specific to IL-23 receptor and IL-12 receptor β 2 subunit.

Single stains were used for compensation controls and fluorescence minus one (FMO) controls were used to help set gates for flow cytometry.

MATERIALS

PBA = phosphate buffered saline or PBS containing 1.5% bovine serum albumin and 0.02% sodium azide

LIVE/DEAD Fixable Aqua dead cell stain purchased from Molecular Probes, *Leiden, The Netherlands*.

Fluorochrome-conjugated monoclonal antibodies:

CD3 (clone REA613, BW264/56), CD4 (REA623), CD8 (BW135/80, REA734), CD14 (REA599, TÜK4), CD16 (REA423), CD25 (4E3), CD45RA (REA562), CD127 (REA614), CD197 (CCR7; REA546), HLA-DR (REA805), IL-12R β 2 (REA333) and IL-23R (218213) (purchased from Miltenyi Biotec, Gladbach Bergische, Germany and R&D Systems, Abingdon, UK).

Optimal quantities to give full staining with minimal amount of each antibody were determined for each vial to ensure optimal results prior to calculation of volume needed.

Cocktails of up to 8 mAbs were made up in PBA buffer.

BD FACS CANTO II flow cytometer

Created: Sep 01, 2019

Last Modified: Oct 29, 2019

PROTOCOL integer ID:
27415

- 1 Acquire blood sample from patient - 3mls EDTA
- 2 Transfer 100mcl whole blood to each FACS tube (7 tubes per patient)
- 3 Add Dead Cell Stain into each tube except tube 7 (unstained), vortex, and incubate for 15 minutes in the dark
- 4 Wash with PBA, vortex, and spin down at 1500RPM for 7 minutes
- 5 Resuspend with appropriate antibody cocktail (50mcl per tube) into each tube except 7 (unstained), vortex, and incubate for 15 minutes
- 6 Add 1ml BD FACS lysing solution, vortex, and incubate for 15 minutes in the dark
- 7 Add 1ml PBA and spin down at 1500RPM for 7 minutes. Discard supernatants
- 8 Resuspend in 2mls PBA, vortex, and spin and discard as per step 6

- 9 Resuspend cells in few drops of PBA, vortex, keep cool, and acquire by flow cytometer immediately