

Feb 20, 2023

Version 2

PIP MRI Magnetization Transfer Ratio (MTR) Measurement of Myelofibrosis in Mouse Tibia V.2

DOI


<https://dx.doi.org/10.17504/protocols.io.5jyl8jwy8g2w/v2>

Thomas L Chenevert¹

¹University of Michigan



Dasha M



Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.5jyl8jwy8g2w/v2>

Protocol Citation: Thomas L Chenevert 2023. PIP MRI Magnetization Transfer Ratio (MTR) Measurement of Myelofibrosis in Mouse Tibia. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.5jyl8jwy8g2w/v2> Version created by **Dasha M**

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

This protocol was tested for 32 diseased mice with 37 test-retest pairs and is used for ongoing murine myelofibrosis model studies

Created: February 17, 2023

Last Modified: February 20, 2023

Protocol Integer ID: 77191

Keywords: preclinical imaging protocol, magnetization transfer ratio, technical repeatability, mouse tibia, myelofibrosis bone marrow, mri measurement of magnetization transfer ratio, pip mri magnetization transfer ratio, tibia bone marrow of myelofibrosis mouse model, mri measurement, mri hardware condition, myelofibrosis mouse model, tibia bone marrow composition in mf mouse model, measurement of myelofibrosis, mtr value, pip mri template, depleting magnetization, mf mouse model tibia, procedures for mtr measurement, mtr measurement, reducing magnetization, myelofibrosis, numerical mtr result, tibia bone marrow composition, tibia bone marrow, mean mtr, mf mouse tibia, bone marrow, mtr, imaging, measured change in the mean mtr, imaging protocol, mf mouse model, resonance saturation rf energy, analysis of mtr map, high mtr

Funders Acknowledgements:

National Institute of Health

Grant ID: U24CA237683

Disclaimer

The PIP MTR claims hold when:

- **Scanner hardware, 3D FLASH acquisition method and parameters, image reconstruction, and data-reduction procedures are equivalent (or superior) to those detailed in section III.**
- **Use of the same animal model and interventions to induce myelofibrosis are performed as detailed in section V.**
- **MTR change is assessed on an individual animal basis where each animal undergoes identical procedures on the same MRI system over longitudinal timepoints.**

Abstract

The goal of this Co-Clinical Imaging Research Program (CIRP) pre-clinical imaging protocol (PIP) is to provide detailed description of key steps used to achieve a stated level of technical repeatability (precision) embodied in "Claims", for MRI measurement of magnetization transfer ratio (MTR) in tibia bone marrow of myelofibrosis mouse models. This pre-clinical imaging procedure document will be referred to as a "profile" and adheres to a PIP MRI template provided in: <https://www.protocols.io/workspaces/pre-clinical-imaging-protocols>.

Here we treat MTR as a "quantitative" metric, although numerical MTR results are a strong function of multiple experimental parameters and MRI hardware conditions. This profile will adopt the common practice to express MTR values in on a 0 to 1 scale. Empirically, MTR reflects fraction of visible signal loss due to application of off-resonance saturation RF energy depleting magnetization in the semi-solid (invisible) macromolecular matrix thus reducing magnetization replenishing visible signal. Simple fluids free of macromolecules should have an MTR $\gg 0$, whereas macromolecule-rich semi-solid tissues such as myelin, muscle, scar and collagen exhibit relatively high MTR (> 0.5). This document details procedures for MTR measurement in MF mouse tibia to achieve stated performance claims.

Tibia bone marrow composition in MF mouse models has gradation going from proximal to distal ends of the tibia, therefore separate claims are made for volume of interest (VOI) analysis of MTR maps for each of three distinct sections along the length of the tibia (see Figure 1):

Section 1 (proximal)° VOI ($\sim 4\text{-}5\text{mm}^3$) within 9mm of proximal end of tibia

Section 2 (transition)° VOI ($\sim 0.4\text{-}0.5\text{mm}^3$) from 10 to 12mm of proximal end of tibia

Section 3 (distal)° VOI ($\sim 0.1\text{-}0.2\text{mm}^3$) from 13 to 14mm of proximal end of tibia

Claim 1: A measured change in the mean MTR in Section 1 VOI of MF mouse model tibia that exceeds ± 0.16 indicates a true biological change has occurred in the tibia bone marrow with 95% confidence.

Claim 2: A measured change in the mean MTR in Section 2 VOI of MF mouse model tibia that exceeds ± 0.11 indicates a true biological change has occurred in the tibia bone marrow with 95% confidence.

Claim 3: A measured change in the mean MTR in Section 3 VOI of MF mouse model tibia that exceeds ± 0.09 indicates a true biological change has occurred in the tibia bone marrow with 95% confidence.

Attachments



UMich_MTR_SOP-

Profil...

738KB

