

Jan 31, 2020

Version 1

Post-IMS Autofluorescence Microscopy V.1

DOI

dx.doi.org/10.17504/protocols.io.879hzh6

Elizabeth Neumann¹, Jamie Allen¹, Heath Patterson¹, Danielle Gutierrez¹, Jeff Spraggins¹

¹Vanderbilt University

VU Biomolecular Multim...

Human BioMolecular Atl...



Jamie Allen

Vanderbilt University

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.879hzh6>

Protocol Citation: Elizabeth Neumann, Jamie Allen, Heath Patterson, Danielle Gutierrez, Jeff Spraggins 2020. Post-IMS Autofluorescence Microscopy. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.879hzh6>

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited



Protocol status: Working

We use this protocol and it's working

Created: November 11, 2019

Last Modified: October 18, 2023

Protocol Integer ID: 29665

Keywords: HuBMAP, Kidney, Microscopy, Autofluorescence, Image Registration, autofluorescence microscopy images of tissue, autofluorescence microscopy image, green autofluorescence microscopy image of the tissue section, microscopy image, green autofluorescence, different imaging modality, correlation of different imaging modality, ims analysis, tissue, image, tissue section

Abstract

Scope:

Obtain autofluorescence microscopy images of tissues after IMS analysis

Expected Outcome:

A brightfield and green autofluorescence microscopy image of the tissue section that enables registration and correlation of different imaging modalities on a pixel by pixel basis.

Troubleshooting

- 1 Place microscope slide within adapter and insert into the Zeiss AxioScan Slide Scanner.
- 2 Perform Coarse focusing of the tissue using:
GFP filter set (ex. 450–490 nm; em. 500–550, green)
High lamp power (~90%) and moderate exposure times (~ 150 ms).
- 3 Perform Fine focusing of the tissue using:
GFP filter set (ex. 450–490 nm; em. 500–550, green)
Brightfield lamp at similar power and exposure time used in step 2 to build a focus map.
- 4 Define the imaging region that includes the tissue.
- 5 Acquire brightfield and autofluorescence image.
- 6 Export autofluorescence image as an OME-TIF File with the following options:
"BigTIFF", "Use Tiles", "Compress", and "Convert to 8 Bit".

Alternatively, other image file types, such as "BigTIFF" and "pngs" can also be useful, depending on the application.