OCT 18, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.8ekhtcw

Protocol Citation: Nus Igem
 2019. Promega BacTiter-Glo Assay. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.8ekhtcw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Oct 18, 2019

Last Modified: Oct 18, 2019

PROTOCOL integer ID:
 28844

🌐 Promega BacTiter-Glo Assay

Nus Igem¹

¹National University of Singapore

Nus Igem

National University of Singapore

MATERIALS

MATERIALS

⊗ BacTiter-Glo Microbial Cell Viability Assay **Promega Catalog #G8230**

- 1 Prepare BacTiter-Glo reagent mix by transferring BacTiter-Glo buffer into BacTiter-Glo substrate

Note

These reagents are part of a set - Promega's BacTiter-Glo Assay.

- 2 Aliquot 50 µL of cells into two wells of 96-well plate (duplicates)
- 3 Add 50 µL of BacTiter Glo reagent mix into each well
- 4 Include PBS-only wells as blank
- 5 Measure luminescence using microplate reader

Equipment

Cytation 5

NAME

Automated microscope; Multi-mode plate reader

TYPE

BioTek

BRAND

N.A.

SKU

<https://www.biotek.com/products/cytation.html>

LINK