VERSION 2

DEC 11, 2020

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bqnbmvn

Protocol Citation: Steven J Burgess 2020. Protein Concentration Determination using Qubit 4 Fluorometer. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bqnbmvn> Version created by Steven J Burgess

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Dec 11, 2020

Last Modified: Dec 11, 2020

PROTOCOL integer ID:
 45475

Protein Concentration Determination using Qubit 4 Fluorometer V.2

Steven J Burgess¹

¹University of Illinois at Urbana-Champaign

Burgess Lab UIUC

Steven J Burgess

University of Illinois at Urbana-Champaign

ABSTRACT

Procedure for quantification of protein concentration using a Qubit 4 Fluorometer. The procedure follows the manufacturer's instructions, this version is adapted for use with samples that have been extracted in protein extraction buffer as part of processing soybean leaf tissue.

Consult the manual for further details:

https://assets.thermofisher.com/TFS-Assets/LSG/manuals/Qubit_Protein_Assay_UG.pdf

IMAGE ATTRIBUTION

Image reproduced from ThermoFisher Scientific website

<https://www.thermofisher.com/us/en/home/industrial/spectroscopy-elemental-isotope-analysis/molecular-spectroscopy/fluorometers/qubit/qubit-fluorometer.html>

MATERIALS

- QubitTM protein assay kit (Life Technologies; [Q33211](#))
- QubitTM assay tubes (Life Technologies; [Q32856](#))
- QubitTM 4 Fluorometer (Life Technologies; [Q33238](#))

BEFORE START INSTRUCTIONS

This protocol assumes protein samples have been resuspended in 1x protein extraction buffer (62.5 mM Tris-HCl (pH 6.8), 2 % SDS (w/v); 10 % glycerol (v/v); 2.5% 2-mercaptoethanol (v/v)) and the dilution factors were determined in relation to processing of 3 size 7 Humboldt leaf disks resuspended in **450 µL** of buffer.

Preliminary tests should be made when working with different samples or volumes

Keywords: protein quantification, qubit, soybean, leaf tissue

Create Working Solution for Analysis

- 1 Create a working solution of Qubit assay buffer by diluting the reagent 1:200 in the provided buffer.

Note

The final volume in each tube must be 200 μL . Each standard tube requires 190 μL of Qubit working solution, and each sample tube requires anywhere from 180–199 μL . Therefore prepare a sufficient Qubit working solution to accommodate all standards and samples.

Create Sample Dilution for Analysis

15m 6s

- 2 Dilute sample 1:50 (196 μL dH_2O + 4 μL sample)

Note

After taking into consideration the Qubit dilution factor (1:20; 10 μL sample + 190 μL qubit assay buffer) the sample being analyzed has been diluted 1:1000. This should give a value in the linear range for qubit (1.25–25 $\mu\text{g}/\text{mL}$).

The dilution is also important to reduce the impact of SDS and 2-mercaptoethanol on quantification. The concentration of SDS must be $>0.2\%$, and 2-mercaptoethanol $>20\text{ mM}$ in the 10 μL sample added to the assay. The concentration of SDS in PEB is 2% , and 2-mercaptoethanol 335 mM, so diluted 1:50 yields [SDS] 0.04% and [2-mercaptoethanol] 6.7 mM, which is in the acceptable range.

- 3 Add 190 μL of Qubit working solution to a fresh Qubit assay tube, one for each sample to be analyzed (including the three protein standards)
- 4 Add 10 μL of protein standard to the appropriate tube and mix by vortexing 00:00:03 s. 3s

5 Add 10 μL of diluted sample to the appropriate tube and mix by vortexing 00:00:03 s. 3s

6 Allow samples to incubate at room temperature for 00:15:00 15m

7 Measure protein sample concentration using the Qubit, following the instructions on the machine (i.e. start by measuring the standard curve).

Note

Remember to adjust measured values according to the dilution factor applied. In this example, after taking into consideration the Qubit dilution factor (1:20; 10 μL sample + 190 μL Qubit assay buffer) the sample has been diluted 1:1000

Equipment

Qubit™ 4 Fluorometer, with WiFi	NAME
Fluorometer	TYPE
Invitrogen	BRAND
Q33238	SKU
https://www.thermofisher.com/order/catalog/product/Q33238#/Q33238	LINK