SEP 26, 2019

Protein Purification strep-tag FPLC

Cleo B.¹

¹Wageningen University

iGEM Wageningen 2019

Cleo B.

ABSTRACT

Protein purification of proteins containing a strep-tag using FPLC.

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.7qchmsw

Protocol Citation: Cleo B. 2019. Protein Purification strep-tag FPLC . **protocols.io** <https://dx.doi.org/10.17504/protocols.io.7qchmsw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Sep 26, 2019

Last Modified: Sep 26, 2019

PROTOCOL integer ID:
 28132

Necessary:

Have performed Protein expression using E. coli strain BL21DE3

To continue see: Protein gel sample preparation V2.

MATERIALS

MATERIALS

MilliQ Water **Contributed by users**

Roche Complete Protease Inhibitor EDTA-Free tablets **Sigma Aldrich Catalog #5056489001**

Buffers

1

1 L Buffer W:

- 100 mM Tris-HCl (pH 8.0)

- 150 mM NaCl
- 1 mM EDTA

Filter this buffer with a 22 um filter

100 mL Buffer E:

- 100 mM Tris-HCl (pH 8.0)
- 150 mM NaCl
- 1 mM EDTA
- 2.5 mM desthiobiotin

1L 0.5 M NaOH (filtered with 22 um filter)

1L MQ (filtered with 22 um filter)

Protein extraction

2 Place post-induction culture from Protein expression using E. coli strain BL21DE3 on ice. For FPLC have 2 L of cell culture.

3 Centrifuge culture at 5000xg for 00:10:00 at 4 °C

1m

4 The pellet is resuspended in 1 mL buffer W per 100 ml cell culture containing one crushed cOmplete mini tablet for 2 L cell culture.

5 The cells are sonicated (VS70 T rod, 25% 1 sec on 2 sec off for 00:05:00 . On ice water). For small amounts use the MS72 rod.

From this point be very sure to keep the cells on ice as much as possible.

6 The cell extract is centrifuged for 00:45:00 at 30 000 g.

4m

7 The supernatant is collected and filtered first with 0.45 μ m filter and then with 0,22 μ m filter. If there are small sample volumes: add 3 mL buffer W.

8

FPLC

9 Connect computer to system

10 Set a manual alarm specific for the column you're using
All speeds are dependent on the machine/column that is used

11 Wash all pumps with MQ

12 Wash the system with MQ (~20 mL)

13 Set valve position to waste

14 Connect column wet (MQ) with a low flow rate (~1 ml/min)
Wash column with MQ

Increase flow rate slowly

- 15** Equilibrate column with buffer W
Slowly increase flow rate
Wait until the conductivity and UV have stabilised
- 16** Wash peristaltic pump with 15 mL MQ
Introduce cell lysate in column using the peristaltic pump (connect wet)
Story flow through and a bit of lysed cells
- 17** Wash the column with buffer W until the lines stabilize
- 18** Start elution with buffer E
Collect all the fractions per 1/1.5 mL
Wash until the protein peak has passed
Place all the fractions on ice immediately
Fractions can be placed on analytical SDS-page gels
- 19** Wash the column with MQ (3 CV)
Wash the column with NaOH (3 CV)
Wash the column with ethanol (decrease speed before changing to ethanol) (3CV)
Wash the system with ethanol