JUL 06, 2021

🌐 qPCR assay for detecting round goby invasive fish species *Neogobius melanostomus*

Omneya Ahmed¹, Alexander Eiler¹

¹eDNA solutions

eDNASolutions

Tech. support phone: +0700264843 email: omneya@ednasolutions.se

Omneya Ahmed Osman

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.bq3umynw

Protocol Citation: Omneya Ahmed, Alexander Eiler 2021. qPCR assay for detecting round goby invasive fish species *Neogobius melanostomus*. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bq3umynw>

MANUSCRIPT CITATION:

Andersen, J.H., Kallenbach, E., Hesselsoe, M., Knudsen, S.W., Møller, P.R., Bekkevold, D., Hansen, B.K & Thaulow, J. (2016): Steps toward nation-wide monitoring of non-indigenous species in Danish marine waters under the Marine Strategy Framework Directive. NIVA Denmark Report. 123 pp. Adrian-Kalchhauser, I., & Burkhardt-Holm, P. (2016): An eDNA assay to monitor a globally invasive fish species from flowing freshwater. PLoS ONE 11: e0147558. doi:10.1371/journal.pone.0147558

DISCLAIMER

DISCLAIMER – FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to protocols.io is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with protocols.io, can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

ABSTRACT

Members of Ponto-Caspian gobies are predatory fish which have colonized freshwaters and brackish waters in Europe and North America causing a lot changes in the native ecosystems.

Several eDNA assays were developed to identify round goby fish. We are going to describe two successful assays.

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Dec 29, 2020

Last Modified: Jul 06, 2021

PROTOCOL integer ID:
45908

GUIDELINES

Handling high concentration of positive controls was performed in a post-PCR room which is physically separated from the pre-PCR room to avoid contamination. Always add your samples first and seal them before adding the serial dilutions of positive control (standard) at the end.

MATERIALS

UltraPure[®]; DEPC-treated Water Thermo Fisher Catalog #10813012

TaqMan[™] Environmental Master Mix 2.0

<https://www.thermofisher.com/order/catalog/product/4396838#/4396838>

Interbal control

http://www.primerdesign.co.uk/assets/files/internal_control_handbook_dna.pdf?timestamp=1504081027

SAFETY WARNINGS

Negative controls of DNase/RNase free water in each qPCR assay.

BEFORE START INSTRUCTIONS

Laboratory work space and equipment were sterilized by UV-light and DNase solution and 70% ethanol. Filter pipet tips were used in all steps of the laboratory work.

DNA extraction

2h 30m

1

A tissue of round goby was extracted with Qiagen DNeasy blood and tissue extraction kit

<https://www.qiagen.com/us/shop/pcr/dneasy-blood-and-tissue-kit/>

The quality of DNA and 260/280 ratio were checked by NanoDrop instrument for nucleic acid measurements.

2

To check for inhibitors in DNA samples, internal control was used,

http://www.primerdesign.co.uk/assets/files/internal_control_handbook_dna.pdf?timestamp=1469446474

OR: <https://www.thermofisher.com/order/catalog/product/4308321>.

3

Primers

A	B	C	D	E	F
Neogobius melanostomus	Primer	Reference	Temperature	Length	GC content
Neo_Mel_COI_F01	5'-CTTCTRGCTCCTCTGGWGTTG-3'	Andersen et al., 2016	59.6 - 62.8	22	54-59
Neo_Mel_COI_R01	5'-CCCWAGAATTGASGARATKCCGG-3'		58.9 - 63.9	23	47-56
Neo_Mel_COI_P01	5'-FAM-CAGGCAACTTRGCACATGCAG-BHQ-3'		60.1 - 62.9	21	52-57
NeoMel_IK_F1	5'-TATGTGATGATCGGACAGC-3'	Adrian-Kalchhauser & Burkhardt-Holm, 2016.		19	53-56
NeoMel_IK_R1	5'-GTTCTCTAGTCAGCTCGCT-3'			19	45-51
NeoMel_IK_Probe	5'-FAM-CATCTTTCTCGGCTTATTCCC-CA-BHQ-3'			23	

4

2 µL Standard DNA dilution

30m

To determine both limit of quantification and limit of detection, genomic DNA of round goby was serially diluted from $1e^2$ - $1e^{-4}$.

5

The composition of PCR mixture

1h 30m

A	B	C	D
PCR reagent	Stock solution	Working solution	Final concentration (µL)
TaqMan Environmental Mastermix 2	2X	1X	10
Forward primer	10 µM	0.4 µM	1
Reverse primer	10 µM	0.4 µM	1
TaqMan probe	2.5 µM	0.1 µM	1
Internal control (IC) primer/probe mix			1
IC-DNA			0.5
DNase/RNase free water			8.5
Template			2

A	B	C	D
Total volume			25

RNase/DNase free water was used as a negative control.

Lower volume of PCR mixture was tested and efficiency was evaluated.

6 - °C Amplification conditions

A	B	C	D
	Step	Time	Temp
50 cycles	Enzyme activation	10 min	95
	Denaturation	15 s	95
	Data collection	1 min	60

Enzyme activation temperature depends on the type of qPCR master mixture.

7 QPCR was performed in BioRad qPCR machine CFX96.

Expected result

Analysis of qPCR results was performed through CFX maestro software

<https://www.bio-rad.com/en-se/product/cfx-maestro-software-for-cfx-real-time-pcr-instruments?ID=OKZP7E15>

8

Expected result

The analysis of PCR internal control

The Cq value obtained with the internal control will vary significantly depending on the extraction efficiency, the quantity of DNA added to the PCR reaction and the individual machine settings. Cq values of 26 or 27±3 are within the normal range.