VERSION 1

DEC 21, 2020

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.bqwymxfw

Protocol Citation: Roey Angel, Eva Petrova, Ana Lara 2020. qPCR: Bacterial SSU rRNA 338F-516P-805R. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bqwymxfw> Version created by Roey Angel

MANUSCRIPT CITATION:

Yu, Y., Lee, C., Kim, J., and Hwang, S. (2005). Group-specific primer and probe sets to detect methanogenic communities using quantitative real-time polymerase chain reaction. *Biotechnol Bioeng* 89, 670–679. doi:10.1002/bit.20347.

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

qPCR: Bacterial SSU rRNA 338F-516P-805R V.1

Roey Angel¹, Eva Petrova¹, Ana Lara¹

¹Soil and Water Research Infrastructure

SoWa RI Anaerobic and Molecular Microbiology (public)

Tech. support email: eva.petrova@bc.cas.cz

Roey Angel

Soil and Water Research Infrastructure

ABSTRACT

Universal 16S rRNA probe-based-qPCR assay for bacteria.

The primers and probe are taken from [Yu et al. \(2005\)](#).

CITATION

Yu Y, Lee C, Kim J, Hwang S (2005). Group-specific primer and probe sets to detect methanogenic communities using quantitative real-time polymerase chain reaction. *Biotechnology and bioengineering*.

LINK

<http://dx.doi.org/10.1002/bit.20347>

MATERIALS

MATERIALS

iQ™ SYBR® Green Supermix **BioRad Sciences Catalog #1708880**

TaqMan™ Fast Advanced Master Mix **Thermo Fisher Scientific Catalog #4444556**

Created: Dec 20, 2020

Last Modified: Dec 21, 2020

PROTOCOL integer ID:
45752

Keywords: qPCR, dual-labelled probe, 16S rRNA gene, bacteria

Primers and probe

1

Name	Type	Sequence	Target region
BAC338F	Forward	ACT CCT ACG GGA GGC AG	338-354
BAC516P ²	Probe	TGC CAG CAG CCG CGG TAA TA	516-536
BAC805R	Reverse	GAC TAC CAG GGT ATC TAA TC	785-805

1. Relative to *E. coli* SSU rRNA gene

2. The probe must be dual-labelled either with 5'-6-FAM, 3'-BHQ1 or any other valid combination

qPCR mixture

2

A	B	C	D
Reagent	Final concentration	1 tube (20 µl)	plate (20 µl x 100)
PCR H2O		2.2	220
2x TaqMan Fast Advanced Master mix	1x	10	1000
BSA (20 µg/µl)	0.4mg/ml	0.4	40
338F	0.5 µM	1.0	100

A	B	C	D
805R	0.5 µM	1.0	100
516P	0.2 µM	0.4	40
Template		5	5 x 100

TaqMan™ Fast Advanced Master Mix **Thermo Fisher Scientific Catalog #4444556**

Thermocycler programme

1h 30m

3

1h 30m

1. 95 °C for 00:05:00
2. x 40 {
 - 2.1 95 °C for 00:00:30
 - 2.2 62 °C for 00:01:00 take snapshot
- }