

Apr 01, 2020

Version 1

Quant-iT™ RiboGreen™ RNA Quantification V.1

DOI

<https://dx.doi.org/10.17504/protocols.io.qdsds6e>

SOWA

Roey Angel¹, Eva Petrova¹

¹Soil and Water Research Infrastructure

Anaerobic and Molecular Microbiology Lab, Biology Centre CAS

Tech. support email: eva.petrova@bc.cas.cz



Roey Angel

Soil and Water Research Infrastructure

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.qdsds6e>

External link: <https://sfamjournals.onlinelibrary.wiley.com/doi/abs/10.1111/1462-2920.12140>



Protocol Citation: Roey Angel, Eva Petrova 2020. Quant-iT™ RiboGreen™ RNA Quantification. **protocols.io**

<https://dx.doi.org/10.17504/protocols.io.qdsds6e>

Manuscript citation:

Angel, R., and Conrad, R. (2013). Elucidating the microbial resuscitation cascade in biological soil crusts following a simulated rain event. Environ Microbiol 15, 2799–2815. doi:10.1111/1462-2920.12140.

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: May 24, 2018

Last Modified: April 20, 2020

Protocol Integer ID: 12434

Keywords: quantification of rna, rna quantification, rna assay kit, rna, ribogreen, using quant, quantification, assay

Abstract

The following protocol is intended for the quantification of RNA using [Quant-iT™ RiboGreen™ RNA Assay Kit](#) (ThermoFisher). This protocol is a simplified and condensed version of the [full protocol](#) from the manufacturer. The procedure described here is for 96 reactions. If samples are run in duplicates, then this should allow quantifying 40 samples.

Attachments



[mp11490.pdf](#)

122KB

Materials

MATERIALS

Quant-iT™ RiboGreen™ RNA Assay Kit **Invitrogen - Thermo Fisher Catalog #R11490**

STEP MATERIALS

Quant-iT™ RiboGreen® RNA Assay Kit **Thermo Scientific Catalog #R11490**

Protocol materials

 Quant-iT™ RiboGreen™ RNA Assay Kit **Invitrogen - Thermo Fisher Catalog #R11490**

 Quant-iT™ RiboGreen® RNA Assay Kit **Thermo Scientific Catalog #R11490**

Safety warnings

 Quant-iT™ RiboGreen® RNA reagent is classified as Not Hazardous. Nevertheless, the user should always consult the MSDS accompanying any of the reagents and apparatus described in this protocol.

Before start

1. This protocol is optimised for measuring an entire 96-well plate. It assumes that 10 wells will be used for measuring the standards and the blank samples (in duplicates) and 86 wells will be used for measuring unknown RNA samples (typically in duplicates).
2. The protocol can be easily adjusted for a lower number of samples by reducing the volume of the working solutions of the reagents. Note though that enough TE should be retained for diluting the standard stock solution (490 or 680 μl), for potentially diluting the unknown samples, if their concentration is too high, and for accounting for pipetting errors. To fill the plate, 19.2 ml of TE is needed. So if only 40 wells are to be used for measuring unknown samples prepare about $50/96 \times 22 \approx 11.5$ ml of TE buffer.
3. The dynamic range of the assay is between 1 ng ml⁻¹ to 50 ng ml⁻¹ in the "low-range" version of the assay and 20 ng ml⁻¹ to 1 $\mu\text{g ml}^{-1}$ in the "high-range" version of the assay. This translates into RNA sample concentrations of 0.2-10 ng μl and 4-200 ng μl^{-1} in the low-range and high-range assays, respectively. Samples with higher RNA concentration need to be diluted (e.g. in RNase-free water or TE buffer).
4. Note that some compounds that often contaminate RNA are known to interfere with the measurement and produce a lower observed measurement. Please refer to the full protocol for a list of these compounds and their effect on the measurement.

Prepare the reaction

- 1 Take out all reagents from the fridge and bring them to room temperature.
Take out the RNA samples from the freezer. RNA samples should be slowly thawed on ice.

20m



Note

Quant-iT™ RiboGreen® RNA reagent is dissolved in dimethylsulfoxide (DMSO), which freezes below 19 °C. The reagent must be completely thawed before using it by bringing it to room temperature. After the reagent thawed, it is advisable to briefly vortex the tube to make sure it is adequately mixed and to spin it down in a centrifuge or a mini centrifuge.

Note

Quant-iT™ RiboGreen® RNA reagent is light sensitive and should be protected from light at all times.

 Quant-iT™ RiboGreen® RNA Assay Kit **Thermo Scientific Catalog #R11490**

- 2 Prepare 22 ml 1X TE buffer by pipetting 1.1 ml of 20X TE buffer into 20.9 ml of nuclease-free water into a sterile and nuclease-free 50 ml tube.
Mix by inverting the tube several times.

2m



 1.1 µL 20x TE buffer

 20.9 µL nuclease-free water

- 3 For high-range quantification:
Dilute the E.coli rRNA standard (100 µg ml⁻¹) to a final concentration of 2 ng µl⁻¹ by mixing 10 µl of rRNA standard–stock with 490 µl 1X TE.

2m

For low-range quantification:

Prepare a 100 ng ml⁻¹ standard by mixing 10 µl of the 2 ng µl⁻¹ work rRNA standard solution with 190 µl 1X TE.

- 4 If needed, prepare a dilution of each sample in 1X TE so that the reading will be within the dynamic range.



Note

It is advisable to run samples in duplicates for a more accurate quantification

5 Prepare the RiboGreen® work solution:

2m

For the **high-range assay** pipette 50 µl Quant-iT™ RiboGreen® RNA Reagent and 9950 µl of 1X TE into a sterile and nucleic-acids free 50 ml tube.

For the **low-range assay** pipette 5 µl Quant-iT™ RiboGreen® RNA Reagent and 9995 µl of 1X TE into a sterile and nucleic-acids free 50 ml tube.

Mix by inverting and keep the solution away from light.

6 Prepare one of the following standard mixtures in the first two columns of a black, sterile, 96-well plate:

10m

Assay version	Diluted RNA std. (either 2 ng µl ⁻¹ or 0.1 ng µl ⁻¹)	1X TE	Final RNA amo unt (ng)
High-range (4-200 ng µl ⁻¹)	100	0	200
Use 2 ng µl ⁻¹ standard	50	50	100
	10	90	20
	2	98	4
	0	0	0
Low-range (200 pg µl - 10 ng µl)	100	0	10
Use 0.1 ng µl ⁻¹ standard	50	50	5
	10	90	1
	2	98	0.2
	0	100	0



Equipment

96-Well microtiter plate

NAME

Polystyrene cell-culture plate

TYPE

Nunc

BRAND

165305

SKU

<https://www.thermofisher.com/order/catalog/product/152028#/152028>^{LINK}

Black, flat-bottom, sterile

SPECIFICATIONS

7 Pipette 99 μ L of TE buffer in the remaining wells.

2m

 99 μ L 1X TE buffer

Note

Tip: use a mechanical or electronic dispenser during this step and step no. 9 to speed up the work.

Equipment

Multipette® E3

NAME

Electronic dispenser

TYPE

eppendorf

BRAND

4987000010

SKU

<http://multipette-system.eppendorf.com/>

LINK



STEP CASE

For low-concentration samples 5 steps

To measure samples with a low concentration, it is possible to reduce the volume of the work solution at this step and pipette more sample in the next step, for a total volume of 100 µl.

8 Pipette 1 µl of the unknown RNA samples in the remaining wells.

5m

 1 µL Unknown RNA sample

Note

Use either a diluted sample in case the concentration is expected to be higher than the dynamic range limit or larger volume in case the concentration is expected to be below the detection limit.

9 Pipette 100 µl of RiboGreen® work solution into each well, including the standard and unknown sample wells

2m

 100 µL RicoGreen work solution

10 Protect the 96-well plate from light and incubate for 2-5 min at room temperature.

5m

 00:05:00

**Measure the samples**

5m

11 Place the plate in a plate reader and measure the fluorescence according to the following parameters:

5m

Excitation	~480 nm
Emission	~520 nm
Integration time	40 s
Lag time	0 s
Gain	Optimal
Number of flashes	10
Calculated well	highest standard
Shaking	5 s



Equipment

Infinite M Nano

Absorbance plate reader

Tecan

TEC006436I

https://lifesciences.tecan.com/plate_readers/fluorescence_absorbance_luminescence?p=tab--2

Note

It is also possible to set the gain to a fixed value (e.g. 100). If the fluorescence values of the standard drop over time this could indicate damage to the reagents or the RNA standard.

- 12 Plot the measured fluorescent values of the standard samples against their known concentrations and fit a linear curve using linear regression. Make sure that the coefficient of determination (R^2) is close to 1 (typically >0.98). Calculate the RNA concentrations in the unknown samples using the slope and intercept parameters of the linear equation. Output values you obtained are in $\text{ng } \mu\text{l}^{-1}$, assuming $1 \mu\text{l}$ of sample was used.

Note

Do not forget to account for any dilutions when calculating the concentration of the RNA in the unknown samples

10m

