NOV 05, 2020

OPEN ACCESS

DOI:
[dx.doi.org/10.17504/protocols.io.bmgck3sw](https://doi.org/10.17504/protocols.io.bmgck3sw)

External link:
<https://doi.org/10.1371/journal.pone.0240823>

Protocol Citation: Yoko Kato-Unoki 2020. Quantification of the molar concentration of the NGS libraries by qPCR. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bmgck3sw>

MANUSCRIPT CITATION:
 Kato-Unoki Y, Umemura K, Tashiro K (2020) Fingerprinting of hatchery haplotypes and acquisition of genetic information by whole-mitogenome sequencing of masu salmon, *Oncorhynchus masou masou*, in the Kase River system, Japan. PLoS ONE 15(11): e0240823. doi: [10.1371/journal.pone.0240823](https://doi.org/10.1371/journal.pone.0240823)

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Quantification of the molar concentration of the NGS libraries by qPCR

PLOS One

Yoko Kato-Unoki

Yoko Kato-Unoki

MATERIALS

MATERIALS

Agencourt AMPure XP **Beckman Coulter Catalog #A63880**

KAPA Library Quantification Kit for Illumina® Platforms **Kapa Biosystems Catalog #KK4835**

Agarose S **Nippon Gene Catalog #312-01193**

BEFORE START INSTRUCTIONS

Note: use low-binding tubes and pipet chips.

Protocol status: Working

Created: Sep 17, 2020

Last Modified: Nov 05, 2020

PROTOCOL integer ID:
42212

- 1 Using the NGS library concentration obtained from the Agilent BioAnalyzer, dilute the libraries to 100 pg/μl with 10 mM Tris-HCl, pH 8.0. Then prepare a 1:10000 dilution of each diluted library (final concentration, 0.01 pg/μl) by conducting 10-fold dilution series using 10 mM Tris-HCl, pH 8.0.

*If the library concentration is unknown, conduct a 10-fold dilution series to achieve an appropriate final dilution (Note: for a 1:3 diluted library, we conducted a 1:500,000 dilution with 10 mM Tris-HCl, pH 8.0).

- 2 Prepare a reaction mix and performs qPCR in accordance with the manufacturer's manual of KAPA Library Quantification kit (KAPA Biosystems).
- 3 Check the product size and presence of adaptor dimers in a 1-μl aliquot of the PCR products by conducting electrophoresis through a 3% agarose–TBE gel. For example, in the Fig. 1, the lanes marked "x" show contamination with adaptor dimer.

Fig. 1. Example agarose electrophoresis.

- 4 Further clean-up the library with Agencourt AMPure XP beads (Beckman Coulter), if the library

contains adaptor dimer. Refer to the Agencourt AMPure XP beads manual for reagent amounts and clean-up procedures. Elute to the same final volume as the original library and re-quantify and check again (i.e, repeat steps 1–3).

- 5 Calculate the molar concentration of each libraries from the standard curve.