VERSION 1

SEP 01, 2021

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.zkff4tn

Protocol Citation: Robyn Hall, Lily Tran, Nina Huang, Ina Smith, Tanja Strive 2021. Rabbit calicivirus capsid Taqman RT-qPCR.

protocols.io

<https://dx.doi.org/10.17504/protocols.io.zkff4tn>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Mar 27, 2019

Last Modified: Sep 01, 2021

PROTOCOL integer ID:
21863

Rabbit calicivirus capsid Taqman RT-qPCR V.1

Robyn
Hall^{1,2},

Lily
Tran¹,

Nina
Huang^{1,2},

Ina
Smith¹,

Tanja
Strive^{1,2}

¹CSIRO Health & Biosecurity; ²Centre for Invasive Species Solutions

Lagovirus biology

Robyn Hall

ABSTRACT

To improve the sensitivity and throughput of molecular testing for rabbit caliciviruses we have developed a multiplex TaqMan RT-qPCR assay that detects the three different pathogenic rabbit calicivirus capsid genotypes currently circulating in Australia. Target sequences for all three variants (Gl.1c, Gl.2, Gl.1a) are located within the VP60 capsid sequence. This will also detect subgenomic RNAs, increasing the sensitivity of the method when compared to assays targeting the non-structural gene sequences. Currently this assay is designed to be run as a one-step RT-qPCR assay for high throughput, rapid diagnostic purposes. It has not been designed for quantification of virus loads. For that purpose, the universal lagovirus SYBR-green assay is still recommended (Hall et al, 2018). This assay has not been designed for the detection of recombinant viruses, and will only return the capsid variant/s present in a sample.

GUIDELINES

To avoid contamination, we recommend that a separate area, consumables, and pipettes are used for reaction setup. All reagents should be handled separately to templates.

MATERIALS

- Gl.1c, Gl.1a, Gl.2 primers and probes - Described in 'Steps'
- SensiFAST Probe no ROX one-step kit Bioline Catalog #BIO-76005
- Neptune fully skirted 96well PCR plate, clear Catalog #3732.x
- Microseal 'B' Adhesive Seals Biorad Sciences Catalog #MSB-1001
- BioRad CFX96 Touch real-time PCR detection system

Keywords: rabbit calicivirus;
diagnostics; Taqman;
multiplex

SAFETY WARNINGS

All work conducted in the laboratory should be undertaken with good laboratory practices in mind. Appropriate personal protective equipment should be worn to protect from biological and chemical hazards. Disposal of waste produced during RNA extraction should be in accordance with local guidelines.

BEFORE START INSTRUCTIONS

Prepare RNA samples for testing. All runs should include a 'no template control' and positive control samples of a GI.1c, GI.1a, and GI.2 RHDV (or a pooled positive control at a minimum).

Plate set-up

1

A	B	C	D	E	F	G	H	I	J	K	L
GI.1	GI.1	RNA 6	RNA 6	RNA 15	RNA 15	RNA 24	RNA 24	RNA 33	RNA 33	RNA 41	RNA 41
GI.2	GI.2	RNA 7	RNA 7	RNA 16	RNA 16	RNA 25	RNA 25	RNA 34	RNA 34	RNA 42	RNA 42
GI.1a	GI.1a	RNA 8	RNA 8	RNA 17	RNA 17	RNA 26	RNA 26	RNA 35	RNA 35	RNA 43	RNA 43
NTC	NTC	RNA 9	RNA 9	RNA 18	RNA 18	RNA 27	RNA 27	RNA 36	RNA 36	RNA 44	RNA 44
RNA 1	RNA 1	RNA 10	RNA 10	RNA 19	RNA 19	RNA 28	RNA 28	RNA 37	RNA 37	RNA 45	RNA 45
RNA 2	RNA 2	RNA 11	RNA 11	RNA 20	RNA 20	RNA 29	RNA 29	RNA 38	RNA 38	RNA 46	RNA 46
RNA 3	RNA 3	RNA 12	RNA 12	RNA 21	RNA 21	RNA 30	RNA 30	RNA 39	RNA 39	RNA 47	RNA 47
RNA 4	RNA 4	RNA 13	RNA 13	RNA 22	RNA 22	RNA 31	RNA 31	RNA 40	RNA 40	RNA 48	RNA 48
RNA 5	RNA 5	RNA 14	RNA 14	RNA 23	RNA 23	RNA 32	RNA 32	RNA 41	RNA 41	RNA 49	RNA 49

Plate layout for rabbit calicivirus capsid Taqman RT-qPCR. Samples and controls are routinely run in duplicate. We run individual controls for each strain; at a minimum a pooled positive control should be run.

Reaction set-up

- In a dedicated Mastermix hood, UV a 96 well qPCR plate and a Microseal B film for 30

2

minutes prior to use.

- Thaw reagents during this time.
- Prepare a 10 μ M primer mix of Gl.1 fwd; Gl.1c_rev; Gl.2 fwd; Gl.2 rv; Gl.1a fwd; Gl.1a rv. Combine 10 μ l of each primer stock (at 100 μ M) with 40 μ l of nuclease-free water.
- Prepare a Mastermix for number of required reactions + 7% extra to account for pipetting inaccuracies:

A	B
	μ l per reaction
Nuclease-free H ₂ O	5
SensiFAST Probe No-ROX One-Step Mix (2x)	7.5
10 μ M primer mix	0.6
10 μ M Gl.1c probe	0.15
10 μ M Gl.2 probe	0.15
10 μ M Gl.1a probe	0.15
RiboSafe RNase Inhibitor	0.15
Reverse transcriptase	0.3
Template RNA	1
TOTAL REACTION VOLUME	15 μ l

Preparation of Mastermix. Include 7% extra to account for pipetting inaccuracies.

A	B	C
Name	Sequence (5' - 3')	Ref
Gl.1 fwd	TTCAGTTYTGGTAYGCCAATGCTG	This study
Gl.1c_rev	AGGCCTGCACAGTCGTAACGTT	Hall et al, 2018
Gl.1c probe	FAM - ATGTCGTTTCGTACCCTTTAACGGCCCTG - BHQ1	This study
Gl.2 fwd	TGGAACCTTGGCTTGAGTGTTGA	Duarte et al, 2015
Gl.2 rv	ACGAGCGTGCTTGTGGACGG	Duarte et al, 2015
Gl.2 probe	HEX - TGTCAGAACTTGTTGAYATYCGCCC - BHQ1	Duarte et al, 2015
Gl.1a fwd	AGTGCAAGTTWTKCTGGGAACAAC	This study
Gl.1a rv	AAGCCTGYACAGTCGTAGCAGC	This study

A	B	C
GI.1a probe	TexasRed - ATGTCATTCTGTGCCYTTTAACAGCCCCA - BHQ2	This study

Primer and probe sets for rabbit calicivirus capsid TaqMan RT-qPCR

- Distribute 14 µl of Mastermix to each well of 96 well PCR plate.
- Add 1 µl of template to respective wells (in duplicate).
 - Ideally, this is done in a dedicated template addition area
- Seal plate with Microseal B film.
- Briefly spin plate down.
- Place plate in qPCR thermocycler (BioRad CFX96) and commence run.

A	B	C
	Temperature (°C)	Duration
Reverse transcription	45	10 min
Initial denaturation	95	2 min
40 cycles of:	95	5 sec
	63	20 sec

Thermocycling conditions for rabbit calicivirus capsid TaqMan RT-qPCR

Analysis

- 3
 - Set baseline to 100 RFU for each fluorophore
 - Quality control:
 1. No amplification in any channel in NTCs.
 2. FAM positive for GI.1 control only (or pooled positive if run).
 3. HEX positive for GI.2 control only (or pooled positive if run).
 4. TexasRed positive for GI.1a control only (or pooled positive if run).
 5. Cq of positive controls within 0.5 Cq of previous runs.
 6. Cq standard deviation <0.5 between replicates.