


Jun 17, 2020

Version 1

SalivaDirect: RNA extraction-free SARS-CoV-2 diagnostics V.1

DOI

<https://dx.doi.org/10.17504/protocols.io.bhfej3je>



Chantal Vogels¹, Doug E. Brackney^{2,1}, Chaney Kalinich¹, Isabel Ott¹, Nathan Grubaugh¹, Anne L Wyllie¹

¹Department of Epidemiology of Microbial Diseases, Yale School of Public Health;

²Department of Environmental Sciences The Connecticut Agricultural Experiment Station

Coronavirus Method De...



Nathan D Grubaugh

Department of Epidemiology of Microbial Diseases, Yale Schoo...

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.bhfej3je>

Protocol Citation: Chantal Vogels, Doug E. Brackney, Chaney Kalinich, Isabel Ott, Nathan Grubaugh, Anne L Wyllie 2020. SalivaDirect: RNA extraction-free SARS-CoV-2 diagnostics. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bhfej3je>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: In development

We are still developing and optimizing this protocol

Created: June 12, 2020

Last Modified: February 11, 2023

Protocol Integer ID: 38086

Keywords: SARS-CoV-2, COVID-19, saliva, extraction-free PCR, COVID-19 diagnostics, saliva samples for sar, saliva for sar, saliva sample, qpcr assay, invalid saliva collection, collecting saliva, rna extraction, saliva of sufficient quantity, salivadirect, saliva viscosity, saliva, rnase, plex assay, rna, qpcr testing, free sar, detection of rp, throughput sar, pcr cycle threshold, negative sar

Abstract

The first section of this protocol details recommendations for collecting saliva for SARS-CoV-2 detection, either by self-collection or healthcare worker-assisted collection.

The second section details an extraction-free method for preparation of saliva samples for SARS-CoV-2 RNA detection. This involves the addition of proteinase K to reduce saliva viscosity, followed by a heating step to inactivate the proteinase K before RT-qPCR testing.

The third section describes a dual-plex RT-qPCR assay for high-throughput SARS-CoV-2 diagnostics. The CDC_N1 (N1) primer/probe set targets the SARS-CoV-2 nucleocapsid gene, and is shown to be sensitive and specific. The human RNase P (RP) primer/probe set is included as an internal sample control - detection of RP at PCR cycle threshold (CT) <35 indicates that saliva of sufficient quantity and quality were tested. The combination of both N1 and RP in the dual-plex assay ensures that any negative SARS-CoV-2 PCR result occurred due to genuine absence of virus, and not because of invalid saliva collection or preparation.

Guidelines

This protocol is still under development.



Materials

MATERIALS

✕ Luna Universal Probe One-Step RT-qPCR Kit - 2,500 rxns **New England Biolabs Catalog #E3006E**

✕ MagMAX[®]; Viral/Pathogen Proteinase K **Thermo Fisher Catalog #A42363**

Sample collection

- Wide-mouth collection tube labelled with unique subject identity number and collection data.
For example: 5 ml screw cap tube, 25 ml conical tube, 50 mL Falcon tube.

Note

15 mL conical tubes are not recommended due to their long, narrow opening. This design increases the risk of contaminating pipettes when aliquoting the collected sample.
Urine cups can also be used for sample collection; however it makes sample vortexing during the processing step difficult and thus we do not recommend using them.

- Personal protective equipment (PPE) for **sample collector**. (at minimum, gloves and face mask)

Extraction-free sample processing

- Thermocycler or real-time PCR (qPCR) instrument
- Vortex mixer
- Plate centrifuge or spinner
- Pipette (P20 and P200)
- Pipette tips (20 µL and 200 µL)
- 96-well PCR plates OR 8-strip PCR tubes (200 µL capacity)
- Adhesive aluminium PCR sealing foil (for PCR plates)
- Proteinase K (50 mg/mL)
- Saliva samples (see "sample collection")

SARS-CoV-2 RNA Detection by Dual-Plex RT-qPCR

- Real-time PCR (qPCR) instrument
- Plate centrifuge or spinner
- 1.5 mL tube centrifuge
- Vortex
- Pipettes (P10, P20, P200, and P1000 - optional P10 multichannel)
- Pipette tips (10 µL, 20 µL, 200 µL, and 1000 µL)
- PCR cooler
- 96-well PCR plates
- 1.5 mL tubes (clear and LightSafe)
- Adhesive transparent PCR seal
- Adhesive film applicator



- **NEB Luna Universal Probe One-Step RT-qPCR kit**
- Nuclease-free water
- RNase away
- 70% ethanol, for cleaning
- Positive control SARS-CoV-2 RNA or **transcripts** (1000 copies/ μ L)
- Extraction-free saliva samples (see "extraction-free sample processing")
- Primers and probes, as follows:

Target	Primer/probe	Sequence
CDC-N1	2019-nCoV_N1-F	GAC CCC AAA ATC AGC GAA AT
	2019-nCoV_N1-R	TCT GGT TAC TGC CAG TTG AAT CTG
	2019-nCoV_N1-P	FAM - ACC CCG CAT TAC GTT TGG TGG ACC - IBFQ
Human RNase P	RP-F	AGA TTT GGA CCT GCG AGC G
	RP-R	GAG CGG CTG TCT CCA CAA GT



		RP-P	Cy5- TTC TGA CCT GAA GGC TCT GCG CG- IBRQ
--	--	------	--

Safety warnings

- ⚠ Processing of any sample type which could potentially be positive for SARS-CoV-2 should be conducted in BSL2+ settings. Before starting work with these samples, please contact your local EHS (environment, health and safety) or biosafety office for proper guidance on how to work with these samples in your laboratory.

Before start

This protocol is still under development. We are currently working on assay optimization, validation of different reagents, and identifying the assay's functional limit of detection.

While collecting saliva is significantly easier than swabs, saliva samples can be difficult to work with. It is important to follow the sample collection guidelines to ensure that saliva, not sputum, is being collected.



- 1 Saliva can either be collected independently by the individual providing the sample ('Self-collection') or with the assistance of a healthcare worker ('Assisted').

STEP CASE

Self-collection

27 steps

For independent collection of sample by the individual providing the sample.

- 2 Before collection, clean hands using alcohol-based sanitizer or soap and water (no fragrances).
- 3 Ensure all collection materials are labelled with the correct identifying information.
- 3.1 While preparing collection materials, begin pooling saliva in your mouth. Saliva production can be stimulated by thinking about food (favorite foods, upcoming meals, etc.) or about the saliva collection itself. Thinking about this while preparing to collect the sample can help ensure enough can be pooled in the mouth quickly and easily.

Note

This protocol is intended for the collection of the **normal** saliva that naturally pools into the mouth. **No coughing or sniffing prior to sample collection is required.** Ideally, water should be avoided 10 minutes prior to collection. Other drinks, food, and nasal sprays should be avoided for half an hour before sample collection.

- 4 Remove the lid of the collection container and gently expel saliva into the collection tube until at least 0.5 mL has been collected.

Note

The total volume measured is to exclude any bubbles.

- 4.1 Once at least 0.5 mL has been collected, securely replace the lid of the collection container.
- 5 Following collection, clean hands using alcohol-based sanitizer or soap and water, and sterilize the collection tube with 70% ethanol or a disinfecting wipe.



- 6 Register the sample collection (including date and time), and place the sample in a secondary container or biohazard bag with a biohazard label.
- 7 Transfer the sample at room temperature to the laboratory for sample processing. The virus RNA in saliva remains stable at room temperature for 3-5 days.
- 8 Store samples at 2-8°C until sample transport or processing (up to 72 hours). If longer-term storage is required, samples can be kept at -20°C for 2-4 weeks, or at -80°C for longer term storage.

Extraction-free sample processing

- 9 Add 2.5 µL of Proteinase K (50 mg/mL) to designated wells of a 96-well PCR plate or 8-strip PCR tubes (200 µL capacity).

Safety information

This work should be completed under BSL-2 conditions, and samples potentially containing SARS-CoV-2 should only be handled in a biosafety cabinet. Please seek guidance from your local biosafety office on specific recommendations for working with samples which could contain SARS-CoV-2.

- 10 Vortex each saliva sample until homogeneous, and immediately transfer 50 µL saliva to each well of the 96-well PCR plate or 8-strip PCR tubes containing proteinase K.
- 10.1 Seal the PCR plate with adhesive aluminium PCR sealing foil, or close the 8-strip tube lids tightly.
- 10.2 Vortex the plate or 8-strip tubes for 1 minute at 3000-5000 RPM.

Note

If using a non-skirted PCR plate or 8-strip tubes, place in a PCR plate rack for stability.

- 10.3 Briefly spin down the plate or tubes using a plate centrifuge or spinner.

Note

If no plate centrifuge or spinner is available, the plate can be gently tapped to get the samples at the bottom of each well.



11 Inactivate the proteinase K by heating samples for 5 minutes at 95°C on a PCR instrument or equivalent thermocycler.

11.1 Briefly spin down the plate or tubes using a plate centrifuge or spinner.

12 Store samples at -80°C or proceed immediately to RT-qPCR testing.

RT-qPCR

13 Prepare 20 uM working stocks of the primers and probes (sequences provided in Materials) by adding 20 µL of 100 uM stock to 80 µL nuclease-free water.

Note

Briefly vortex and centrifuge reagents before use.
Probes are photosensitive and should be stored in the dark.

13.1 Use the 20 uM working stocks to prepare Multiplex primer-probe-water mix containing the following:

Component	Volume (1 reaction)	Volume (100 reactions)
2019-nCoV_N1-F (400 nM/reaction)	0.4 µL	40 µL
2019-nCoV_N1-R (400 nM/reaction)	0.4 µL	40 µL
2019-nCoV_N1-P (200 nM/reaction)	0.2 µL	20 µL
RP-F (150 nM/reaction)	0.15 µL	15 µL
RP-R (150 nM/reaction)	0.15 µL	15 µL



RP-P (200 nM/reaction)	0.2 µL	20 µL
Nuclease-free water	2.5 µL	250 µL

A larger volume of primer-probe-water mix can be prepared in advance, aliquoted in LightSafe microcentrifuge tubes, and stored at -20°C.

- 14 On ice, prepare a master mix containing the following (account for 10% extra lost during pipetting).

Note

Briefly vortex and centrifuge reagents before use.

Component	Volume per reaction
Master mix	10 µL
RT	1 µL
Primer-probe-water mix	4 µL

5 µL Proteinase K-treated samples, standards, or controls will be added to each well individually, for a total 20 µL reaction.

- 14.1 Place the 96-well PCR plate on the PCR cooler and add 15 µL of mastermix to each designated well.

- 15 Wipe the aluminum seal of the processed samples plate with RNase away and 70% ethanol.



Safety information

Many institutions will require samples potentially containing full-length SARS-CoV-2 RNA to be handled in a biosafety cabinet. Please seek guidance from your local biosafety office on specific recommendations for working with samples which could contain full-length SARS-CoV-2 RNA.

- 15.1 Carefully pierce through the foil seal and add 5 μ L of extraction-free saliva sample to each well of the mastermix plate. Mix by pipetting, taking care to avoid introducing bubbles.

Note

Extraction-free saliva samples can be added using a multichannel pipette for high-throughput testing.

RT-qPCR

- 15.2 Add 5 μ L of positive control (e.g. SARS-CoV-2 RNA or RNA transcript at $\sim 10^3$ RNA copies/ μ L) and no-template control (NTC - water) to each designated PCR tube. Mix by pipetting, taking care to avoid introducing bubbles.
- 15.3 Seal with a transparent plastic qPCR seal. Centrifuge briefly to remove bubbles, if present.
- 16 Load the plate into the qPCR machine, and run the following thermocycler conditions:

	Step	Temperature	Time
	1	55°C	10 min
	2	95°C	1 min
	3	95°C	10 sec
	4	55°C	30 sec
	5	Read plate	



Repeat steps 3-5 for 40 cycles.

17 Report results per the following criteria:

	Output	Significance	RP CT	N1 CT
	0	Negative	<35	>40
	1	Invalid	>35	>40
	2	Positive	any value	<40