

Jun 25, 2025

Version 2

Saturation Recovery V.2



In 1 collection

DOI

<https://dx.doi.org/10.17504/protocols.io.kxygxywbkl8j/v2>

Alexander L. Paterson¹

¹National Magnetic Resonance Facility at Madison (NMRFAM), University of Wisconsin-Madison, Madison, WI, United States



NMRFAM Facility

UW-Madison

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.kxygxywbkl8j/v2>

Protocol Citation: Alexander L. Paterson 2025. Saturation Recovery. **protocols.io**

<https://dx.doi.org/10.17504/protocols.io.kxygxywbkl8j/v2> Version created by **NMRFAM Facility**

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited



Protocol status: Working

We use this protocol and it's working

Created: June 25, 2025

Last Modified: June 25, 2025

Protocol Integer ID: 221035

Keywords: Measuring Spin-1/2 T1 via Saturation Recovery, saturation recovery purpose, longitudinal relaxation time, resonances present in the sample, recycle delay, recycle delays appropriate for the purpose, resonance, initial 1d experiment, later experiment, completion of this sop, recovery, sample

Funders Acknowledgements:

National Science Foundation

Grant ID: 1946970

Abstract

Purpose

To measure the longitudinal relaxation time, T_1 , of resonances present in the sample.

Scope

This SOP should be completed after an initial 1D experiment has been completed, but before further experiments are performed. Upon completion of this SOP, the user will be able to set recycle delays appropriate for the purpose of later experiments, either to maximize signal per unit time, or to ensure quantitative intensities.

Guidelines

If acquisition of a 1D spectrum requires substantial signal averaging, this SOP may not be appropriate.

This SOP is written for acquisition using a Bruker console and TopSpin 3.6.2. It may be adapted to other TopSpin versions where appropriate. The basic process is the same on TopSpin 4.x, but requires a NEO-compatible pulse sequence. The sequence **satrect1.neo.nmrfam** is compatible.

This SOP is intended for a beginning user who has been trained on basic spectrometer safety and principles.

Materials

Definitions:

1. T_1 : Longitudinal relaxation time
2. I20: Number of saturation pulse loops
3. d20: Presaturation loop delay

Appendix:

If the presaturation train is not completely effective, the recovery curve might initially decrease, and then follow the expected exponential curve. This is common behaviour and can be dealt with by discarding the initial points of the curve.

The values of the variable delay list T1_standard_array range from 0.25 to 128 s, increasing by a factor of two for each increment. This should provide adequate spacing to accurately determine T_1 values between 1 s and 64 s. If necessary, adjust the list for shorter or longer values of T_1 .




Figure 1. Example of a successful ^7Li saturation recovery experiment performed on a lithium silicate glass. The T_1 of this environment was found to be 3.4 s.



Safety warnings

- ❗ The minimum length of **d1** for the 2D experiment will depend on whether decoupling is used during acquisition. If decoupling is used, **d1** should be about 20 times as long as the acquisition period, typically on the order of 1 s. If decoupling is not used, a **d1** value of 100 ms may be reasonable.

Before start

User is responsible for knowing the duty cycle of the probe and instrument being used.

User is responsible for knowing the high-power decoupling limits of the probe.

Expected time to completion: 1 hour or less

Procedure

- 1 Prepare the experimental environment.
 - 1.1 Create a new experiment window.
 - 1.2 Load the **satrect1** pulse program.
 - 1.3 Set the observe nucleus pulse length, **p1**, and pulse power, **pl1**, to the previously measured pulse parameters.
 - 1.4 If necessary, enable decoupling by setting the -Ddec flag in zgoptions.
 1. Set the ¹H decoupling program, **CPDPRG2**, pulse length, **PCPD2**, and pulse power, **plw12**, to a previously optimized value.
 - 1.5 Set the recycle delay, **d1**, to a relatively long value, *e.g.*, 10 s.
 - 1.6 Set the number of saturation pulse loops, **l20**, to 0.
 - 1.7 Set the variable delay list, **vdlist**, to T1_standard_array.
- 2 Acquire an initial experiment. The number of scans should be as small as practicable to allow for adequate signal, ideally as low as 1.
 - 2.1 Ensure that all expected resonances are visible. If some are not visible, either reduce the pulse length or increase **d1**, as appropriate.
- 3 Set the number of saturation pulse loops, **l20**, to an initial value of 10. Set the presaturation loop delay, **d20**, to an initial value of 100 μs.
- 4 Acquire a spectrum. There should be no signal present.



- 4.1 If signal is still present, increase **I20**, up to a maximum of 100, and reacquire the spectrum.
- 4.2 If increasing **I20** does not result in a null signal, reduce the number of loops and retry with longer presaturation loop delays **d20**.
- 4.3 If increasing **d20** does not result in a null signal, retry with shorter presaturation loop delays **d20**.
- 5 Convert the experiment to a 2D experiment window.
- 5.1 Move to the acqupars window via the command **eda**.
- 5.2 Click on the button titled "Change acquisition dimension of current data set" and change the dimension to 2D.
- 5.3 Set the following parameters:
 1. **F1 TD**: 12
 2. **FnMODE**: QF
 3. **F1 SI**: 16
 4. **d1**: A minimal value that respects the probe duty cycle, see Warnings.
- 6 Acquire the 2D experiment via the command **zg**.
- 7 Process the 2D experiment.
- 7.1 Transform the data using the command **xf2**.
- 7.2 Correct the F2 phasing of the slice with the longest delay period.
- 7.3 Do not transform or phase the F1 axis.
- 8 To process the data, launch the TopSpin utility **t1guide** and follow the workflow in the utility.



- 8.1 Extract the slice with the longest variable delay, *i.e.*, the most recently acquired slice.
- 8.2 Integrate each resonance. Save using the option "Export Regions to Relaxation Module and .ret".
 1. If resonances are not resolved, attempt to select the regions of least overlap.
 2. If resonances cannot be reasonably separated, integrating over all of the peaks will allow for the determination of the longest T_1 component.
- 8.3 Enter the relaxation window and set the Fitting Function Function Type to **uxnmrt1**.
- 8.4 Calculate fit for all peaks. If successful, the fitting window should resemble the example in the appendix.
- 9 Record the T_1 for each site. Use the longest value of T_1 for determining the recycle delay of future experiments.

Protocol references

Bruker Solid State NMR User Manual Z4D10641B, Section 16.2.4

Y. Nishiyama and N. T. Duong, "Practical guides for ^1H detected solid-state NMR under fast MAS for small molecules," Journal of Magnetic Resonance Open, vol. 10-11, p. 100062, Jun. 2022.

<https://doi.org/10.1016/j.jmro.2022.100062>

Example Data at 1.1 GHz: usnan.org/ark:/83454/e1f510d55c-7c34-4a84-a88d-7eb37a629b3c

Example Data at 900 MHz: usnan.org/ark:/83454/e11dd43f92-1862-49ed-b56a-77e84adf9366

Example Data at 600 MHz: usnan.org/ark:/83454/e138cef24e-755f-4e60-ab44-0f487bb3a268