AUG 10, 2019

OPEN ACCESS

DOI:
[dx.doi.org/10.17504/protocols.io.6a8hahw](https://doi.org/10.17504/protocols.io.6a8hahw)

External link:
<https://www.biorxiv.org/content/10.1101/607242v2>

Protocol Citation: Charles C Horn, Derek M. Miller, Stephanie Fulton, Bill J. Yates, Lee E. Fisher, Ameya C. Nanivadekar 2019. SPARC - Gastrointestinal myoelectric recordings from the behaving ferret. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.6a8hahw>

MANUSCRIPT CITATION:
[Machine learning prediction of emesis and gastrointestinal state in ferrets Ameya C. Nanivadekar, Derek M. Miller, Stephanie Fulton, Liane Wong, John Ogren, Girish Chitnis, Bryan McLaughlin, Shuyan Zhai, Lee E. Fisher, Bill J. Yates, Charles C. Horn bioRxiv 607242; doi: <https://doi.org/10.1101/607242>](#)

SPARC - Gastrointestinal myoelectric recordings from the behaving ferret

PLOS One

Charles C Horn¹, Derek M. Miller¹, Stephanie Fulton¹, Bill J. Yates¹, Lee E. Fisher¹, Ameya C. Nanivadekar¹

¹University of Pittsburgh

SPARC

Tech. support email: info@neuinfo.org

Charles C Horn
University of Pittsburgh

ABSTRACT

This protocol provides the steps for recording the gastrointestinal myoelectric responses and stimulation of the abdominal vagus nerve from the behaving ferret. These tests occur after an animal is surgically implanted with gastrointestinal and vagus nerve electrodes (see the protocol "SPARC - Chronic implantation of gastrointestinal and vagus nerve electrodes in the ferret" for implantation methods).

Funding: This protocol was developed with funding from the NIH Common Fund's Stimulating Peripheral Activity to Relieve Conditions (SPARC) program (Award U18TR002205). To learn more about the SPARC program, visit <https://sparc.science>.

GUIDELINES

The following guidelines are important:

- 1) care must be taken to not damage the head connector (e.g., getting debris inside the connection)
- 2) a ferret will often play with the tether, especially during initial exposure; therefore, the position of the tether will need to be manipulated manually to avoid the animal getting tangled

MATERIALS

MATERIALS

☒ feeding tube **Bio-Serv Catalog #9012**

☒ Emetine dihydrochloride hydrate **Sigma-aldrich Catalog #45160**

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Aug 08, 2019

Last Modified: Aug 10, 2019

PROTOCOL integer ID:
26688

Keywords: gastrointestinal tract, stomach, intestine, vagus nerve, electrophysiology, behavior

Room setup

- 1

Cover floor of behavioral test chamber with layer of Alpha-dri bedding. The floor of the plexiglas chamber is 51 x 51 cm (<https://www.ncbi.nlm.nih.gov/pmc/articles/PMC3348962/>)
- 2

Turn off all lights except lamp over testing chamber.
- 3

Set up video camera directly above chamber so that entire area is visible with minimal shadows. Set up a second camera to view the feeding tube from the side for feeding trials to track volume consumed.

Equipment

video camera	NAME
camera	TYPE
Sony	BRAND
HDR-XR550V	SKU

- 4 Suspend tether ~50 cm above chamber and secure with rubber bands to provide elasticity as the animal moves.
- 5 Keep the room quiet during recording except for white noise.

Animal

- 6 Animal should be fasted 3 hours before starting recording.
- 7 Place animal in chamber and allow 10 minutes for acclimation.
- 8 Connect animal to tether.
- 9 During testing, hold the tether straight above the animal out of its reach, but with some slack. If the animal tries to grab the tether, especially during adaptation, a plastic straw can be inserted

through the breathing holes in the side of the chamber to distract the animal from the tether.

Recording: Baseline control

- 10 Record electrode impedances before the recording begins.

Equipment

Grapevine Neural Interface Processor and a Nano2 recording headstage	NAME
electrophysiological recording and stimulation	TYPE
Ripple LLC	BRAND
Grapevine Neural Interface Processor and a Nano2 r	SKU

- 11 Electrophysiological recording and video recording should be started at the same time.

Equipment

Media Recorder	NAME
video to computer recording system	TYPE
Noldus Information Technology	BRAND
MediaRecorder	SKU
https://www.noldus.com/human-behavior-research/products/mediarecorder	LINK

- 12 Record 1 h of baseline data for control days.

Recording: Feeding

- 13** For feeding trials, introduce Ensure liquid diet in a graduated tube (100 ml Richter glass feeding tube) attached to the side of the chamber at 10 min after the start of recording. Remove the feeding tube 45 min after presentation. Finish with 10 min of baseline recording.

Equipment

Grapevine Neural Interface Processor and a Nano2 recording headstage	NAME
electrophysiological recording and stimulation	TYPE
Ripple LLC	BRAND
Grapevine Neural Interface Processor and a Nano2 r	SKU

Recording: Feeding with vagus stimulation

- 14** For feeding with stimulation trials, introduce Ensure at the same time as in Step 13. Start stimulation 2 minutes after food introduction. Stop stimulation and remove food at the same time, i.e., minute 45 in the recording. Finish with 10 minutes of baseline

Equipment

Grapevine Neural Interface Processor and a Nano2 recording headstage	NAME
electrophysiological recording and stimulation	TYPE
Ripple LLC	BRAND
Grapevine Neural Interface Processor and a Nano2 r	SKU

Recording: Emetine (emesis testing)

- 15** For emetine trials, begin infusion at the 10 min in the recording. Infusion rate is 10ml/min. Record for 1 hour after infusion has ended.

Equipment

Grapevine Neural Interface Processor and a Nano2 recording headstage	NAME
electrophysiological recording and stimulation	TYPE
Ripple LLC	BRAND
Grapevine Neural Interface Processor and a Nano2 r	SKU

- 15.1** After recording is done, animal should be monitored until there is no further emesis.

Recording: Water control

- 16** For water control trials, follow Step 15.

Equipment

Grapevine Neural Interface Processor and a Nano2 recording headstage	NAME
electrophysiological recording and stimulation	TYPE
Ripple LLC	BRAND
Grapevine Neural Interface Processor and a Nano2 r	SKU

End

- 17** Stop electrophysiological and video recordings at the same time.

- 18** Measure impedances at the end of each recording.

Equipment

Grapevine Neural Interface Processor and a Nano2 recording headstage	NAME
electrophysiological recording and stimulation	TYPE
Ripple LLC	BRAND
Grapevine Neural Interface Processor and a Nano2 r	SKU