

Nov 07, 2017

## Standard hematological analyses HIPEC

 PLOS One

DOI

[dx.doi.org/10.17504/protocols.io.ki6cuhe](https://doi.org/10.17504/protocols.io.ki6cuhe)

Sven Van Poucke<sup>1</sup>, Dana Huskens<sup>2</sup>, Kurt Van der Speeten<sup>3</sup>, Mark Roest<sup>2</sup>, Bart Lauwereins<sup>4</sup>, Ming-Hua Zheng<sup>5</sup>, Seppe Dehaene<sup>4</sup>, Joris Penders<sup>6</sup>, Abraham Marcus<sup>7</sup>, Marcus Lancé<sup>7</sup>

<sup>1</sup>Department of Anesthesiology, Intensive Care, Emergency Medicine and Pain Therapy, ZOL, Genk, Belgium;

<sup>2</sup>Synapse Research Institute, Maastricht, The Netherlands.; <sup>3</sup>Department of Surgery, ZOL, Genk, Belgium.;

<sup>4</sup>Department of Anesthesiology, Intensive Care, Emergency Medicine and Pain Therapy, ZOL, Genk, Belgium.;

<sup>5</sup>Department of Hepatology, the First Affiliated Hospital of Wenzhou Medical University, Wenzhou 325000,

China.;

<sup>6</sup>Central Diagnostic Laboratory, ZOL, Genk, Belgium.;

<sup>7</sup>Department of Anesthesiology, ICU and Perioperative Medicine, HMC, Doha, Qatar.

HIPACAT



Sven Van Poucke

Department of Anesthesiology, Intensive Care, Emergency Medi...

OPEN  ACCESS



DOI: [dx.doi.org/10.17504/protocols.io.ki6cuhe](https://doi.org/10.17504/protocols.io.ki6cuhe)

External link: <https://doi.org/10.1371/journal.pone.0193657>

**Protocol Citation:** Sven Van Poucke, Dana Huskens, Kurt Van der Speeten, Mark Roest, Bart Lauwereins, Ming-Hua Zheng, Seppe Dehaene, Joris Penders, Abraham Marcus, Marcus Lancé 2017. Standard hematological analyses HIPEC . **protocols.io** <https://dx.doi.org/10.17504/protocols.io.ki6cuhe>

**Manuscript citation:**

Poucke SV, Huskens D, Speeten KVd, Roest M, Lauwereins B, Zheng M, Dehaene S, Penders J, Marcus A, Lancé M (2018) Thrombin generation and platelet activation in cytoreductive surgery combined with hyperthermic intraperitoneal chemotherapy - A prospective cohort study. PLoS ONE 13(6): e0193657. doi: [10.1371/journal.pone.0193657](https://doi.org/10.1371/journal.pone.0193657)

**License:** This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

**Protocol status:** Working

**Created:** November 01, 2017

**Last Modified:** March 08, 2018

**Protocol Integer ID:** 8510

**Keywords:** HIPEC

## Abstract

Standard hematological analyses

EDTA-anticoagulated blood was used for cytometric analysis using a whole blood counter Sysmex XE 2100® (Sysmex, Kobe, Japan) to obtain a whole blood count. Fibrinogen levels were determined with an ACL-9000 (Diamond Diagnostics, Holliston, MA) coagulation analyser, using a PT-fibrinogen high sensitivity reagent, which is a high-sensitivity calcium thromboplastin that allows the simultaneous determination of PT and fibrinogen levels. aPTT was measured using an ACL-9000 coagulation analyser and INR was calculated by the formula INR (PT patient/PT normal).

