AUG 03, 2018

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.sbweape

External link:
<https://www.liebertpub.com/doi/10.1089/3dp.2018.0060>

Protocol Citation: Anton Du Plessis 2018. Standard method for microCT-based additive manufacturing quality control 3: surface roughness. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.sbweape>

MANUSCRIPT CITATION:
A du Plessis, I Yadroistava, SG le Roux, I Yadroitsev, 'X-ray microcomputed tomography in additive manufacturing: a review of the current technology and applications', 3D printing and additive manufacturing. <http://doi.org/10.1089/3dp.2018.0060>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Standard method for microCT-based additive manufacturing quality control 3: surface roughness

Anton Du Plessis¹

¹University of Stellenbosch

Anton Du Plessis

ABSTRACT

The use of microCT of 10 mm coupon samples produced by AM has the potential to provide useful information of mean density and detailed porosity information of the interior of the samples. In addition, the same scan data can be used to provide surface roughness analysis of the as-built surfaces of the same coupon samples. This can be used to compare process parameters or new materials. While surface roughness is traditionally done using tactile probes or with non-contact interferometric techniques, the complex surfaces in AM are sometimes difficult to access and may be very rough, with undercuts and may be difficult to accurately measure using traditional techniques which are meant for smoother surfaces. This standard workflow demonstrates on a coupon sample how to acquire surface roughness results, and compares the results from roughly the same area of the same sample with tactile probe results. The same principle can be applied to more complex parts, keeping in mind the resolution limit vs sample size of microCT.

ATTACHMENTS

[Method for 10 mm cube surface roughness measurement.pdf](#)

[03 Cube surface roughness Edit.mkv](#)

Protocol status: Working
We use this protocol and it's working

Created: Aug 03, 2018

Last Modified: Aug 03, 2018

PROTOCOL integer ID:
14422

Keywords: additive manufacturing, microCT, x-ray tomography, surface roughness, quality control