FEB 26, 2019

Standardized immunohistochemical staining used in the Human Protein Atlas

Anna Martinez Casals^{1,2}, Cecilia Lindskog³

¹Science for life laboratory; ²KTH - Royal Institute of Technology;

³Science for Life Lab / Science for life laboratory / KTH - Royal Institute of Technology

Human Protein Atlas

Tech. support email: contact@proteinatlas.org

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.yj8furw

External link:

<https://www.ncbi.nlm.nih.gov/pubmed/22688270>

Protocol Citation: Anna Martinez Casals, Cecilia Lindskog 2019. Standardized immunohistochemical staining used in the Human Protein Atlas. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.yj8furw>

MANUSCRIPT CITATION:

Production of tissue microarrays, immunohistochemistry staining and digitalization within the human protein atlas. Kampf C, Olsson I, Ryberg U, Sjöstedt E, Pontén F. J Vis Exp. 2012 May 31; (63). pii: 3620. doi: 10.3791/3620.

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Anna Martinez Casals

Science for life laboratory ; KTH - Royal Institute of Techn...

ABSTRACT

The Human Protein Atlas provides a map showing the distribution and relative abundance of proteins in the human body. All IHC staining in the Human Protein Atlas project are performed using the following standard protocol. The primary antibody dilution is based on titration optimization, the dilution suggested by the Human Protein Atlas can be found under antibody and antigen information for each antibody. When primary antibody originates from other host animals than rabbit, there are some modifications and different secondary antibody is used.

Protocol status: Working
We use this protocol in our group and it is working.

Created: Feb 26, 2019

Last Modified: Feb 26, 2019

PROTOCOL integer ID:
20832

MATERIALS

MATERIALS

- ⊗ DAB quanto substrate system **Thermo Fisher Scientific Catalog #Cat# TA-125-QHDX**
- ⊗ UltraVision LP HRP polymer **Thermo Fisher Scientific Catalog #Cat# TL-12HL**
- ⊗ Ultra V Block **Thermo Scientific Catalog #cat# TA-060-UB**
- ⊗ Mayer's hematoxylin plus **Histolab Catalog #cat# 01825**
- ⊗ Hydrogen peroxid 30% **Merck Millipore Catalog #1.07209.1000**
- ⊗ NeoClear **VWR International Catalog # 1.09843.5000**
- ⊗ Lithium carbonate **Merck Millipore Catalog #1.05680.0250**
- ⊗ Pertex **Histolab Catalog #00871.0500**
- ⊗ Antibody Diluent OP Quanto **Thermo Fisher Scientific Catalog #TA-125-AD**
- ⊗ Tris Buffered Saline & Tween 20 (20x) **Thermo Fisher Scientific Catalog #TA-999-TT**
- ⊗ Lab Vision™ Tween™ 20 Detergent **Thermo Fisher Scientific Catalog #cat# TA-125-TW**
- ⊗ Coverslips 24x50mm **VWR International Catalog #631-0146**
- ⊗ Thermo Scientific™ SuperFrost Plus™ Adhesion slides **Thermo Fisher Scientific Catalog #J1800AMNZ**
- ⊗ Lab Vision™ PT Module™ Deparaffinization and Heat-Induced Epitope Retrieval Solutions (100X) **Thermo Fisher Scientific Catalog #TA-250-Pm1x**

Wash buffer

9.5L distilled water
500ml Tris Buffered Saline & Tween 20 (20x)
15ml Large Volume Tween 20

Retrieval buffer

5L distilled water
50ml PT Module Buffer 1

SAFETY WARNINGS

⚠ Regulations about working with tissue samples may vary between institutions, it is important to be aware about the guidelines before to start any experiment.

3,3'-Diaminobenzidine (DAB) is toxic if swallowed, in contact with skin or if inhaled. It may cause cancer and damage to organs.

BEFORE START INSTRUCTIONS

Cut Formalin-Fixed Paraffin-Embedded (FFPE) tissue specimen at 4 µm thickness using a water fall microtome:

HM 355S Automatic Microtome, ThermoFisher Scientific, 905200
Section Transfer System (STS), ThermoFisher Scientific, 771200

Place the section on a superfrost glass slide.

The time the sections can be left in the water depends on the type of paraffin waxed used and the water temperature.

The lab uses paraffin wax from HistoLab Products AB, which have a melting point of 56-58 °C and we recommend a water bath temperature of 37-39 °C.

Deparaffinization

- 1 Dry paraffin sections at room temperature overnight.
- 2 Bake the paraffin sections from ⌚ 12:00:00 to ⌚ 24:00:00 at 🌡 50 °C .
- 3 Xylene incubation: incubate slides in xylene for ⌚ 00:05:00 .

incubate slides in xylene for 00:05:00 .

incubate slides in xylene for 00:01:00 .

- 4 Ethanol absolute incubation: incubate slides in ethanol absolute for 00:03:00 .
incubate slides in ethanol absolute for 00:03:00 .

- 5 96% ethanol incubation & 30%H2O2 (1:100): incubate slides for 00:05:00 .

- 6 96% ethanol incubation: incubate slides in 96% ethanol incubation for 03:00:00 .

- 7 80% ethanol incubation: incubate slides in 80% ethanol incubation for 00:03:00 .

- 8 Distilled water: : incubate slides in distilled water or until antigen retrieval step.

Standard antigen retrieval method

- 9 Heat the slides immersed in retrieval buffer for 00:04:00 at 125 °C in a pressure boiler (Decloaking chamber model DC2008INTL).
- 10 After completed boiling, leave the slides in the pressure boiler and let them cool down till 90 °C .

Total time 00:45:00

Immunohistochemical staining program, Autostainer 480

- 11 Place slides in the Autostainer 480, all the incubations are done at room temperature.

Equipment

Autostainer 480	NAME
ThermoFisher Scientific	BRAND
cat# A80500007	SKU
https://www.thermofisher.com/order/catalog/product/A80500008	LINK

- 12 Rinse slides in wash buffer.

- 13 Incubate slides with Ultra V Block for 00:05:00 .

- 14 Rinse slides in wash buffer go to step #14 once more, total washes 2 .

- 15 Incubate slides with primary antibody, diluted in antibody Diluent OP, for 00:30:00 .

- 16 Rinse slides in wash buffer go to step #16 two times more, total washes 3 .
- 17 Incubate slides with labeled HRP polymer for 00:30:00 .
- 18 Rinse slides in wash buffer go to step #18 once more, total washes 2 .
- 19 Incubate slides with DAB solution for 00:05:00 .
- 20 Rinse slides in distilled water.

Counterstaining and coverslipping

- 21 Transfer slides in the Autostainer XL.

Equipment

Autostainer XL

NAME

Leica biosystems

BRAND

cat# Leica ST5010 Autostainer XL

SKU

<https://www.leicabiosystems.com/histology-equipment/he-stainers-special-stainers-coverslippers/products/leica-st5010/>

LINK

- 22 Counterstain slides with hematoxylin for 00:07:50 .
- 23 Rinse slides in lithium carbonate water diluted 1:5 from saturated solution for 00:01:00 .
- 24 Rinse slides in tap water for 00:05:00 .
- 25 80% ethanol incubation: incubate slides in 80% ethanol for 00:03:00 .
- 26 96% ethanol incubation: incubate slides in 96% ethanol for 00:03:00 .
incubate slides in 96% ethanol for 00:03:00 .
- 27 99% ethanol incubation: incubate slides in 99% ethanol for 00:03:00 .

incubate slides in 99% ethanol for 00:03:00 .

incubate slides in 99% ethanol for 00:03:00 .

28 NeoClear incubation: incubate slides in NeoClear for 00:03:00 .

incubate slides in NeoClear for 00:03:00 .

29 Mount coverslip in each of the slide using Pertex as a mounting media.