VERSION 1

FEB 26, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.yjffukn

Protocol Citation: Anna Martinez Casals, Cecilia Lindskog 2019. Standardized immunohistochemical staining used in the Human Protein Atlas. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.yjffukn>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol in our group and it is working.

Created: Feb 25, 2019

Last Modified: Feb 26, 2019

PROTOCOL integer ID:
20811

Standardized immunohistochemical staining used in the Human Protein Atlas V.1

Anna Martinez Casals^{1,2}, Cecilia Lindskog³

¹Science for life laboratory; ²KTH - Royal Institute of Technology;

³Science for life laboratory / KTH - Royal Institute of Technology

Human Protein Atlas

Tech. support email: contact@proteinatlas.org

Anna Martinez Casals

Science for life laboratory ; KTH - Royal Institute of Techn...

ABSTRACT

All IHC staining in the Human Protein Atlas project are performed using the following standard protocol. The primary antibody dilution is based on titration optimization, the dilution suggested by the Human Protein Atlas can be found under antibody and antigen information for each antibody. When primary antibody originates from other host animals than rabbit, there are some modifications and different secondary antibody is used.

MATERIALS

MATERIALS

DAB quanto substrate system **Thermo Fisher Scientific Catalog #Cat# TA-125-QHDX**

UltraVision LP HRP polymer **Thermo Fisher Scientific Catalog #Cat# TL-12HL**

Ultra V Block **Thermo Scientific Catalog #cat# TA-060-UB**

Mayer's hematoxylin plus **Histolab Catalog #cat# 01825**

SAFETY WARNINGS

Regulations about working with tissue samples may vary between institutions, it is important to be aware about the guidelines before to start any experiment.

3,3'-Diaminobenzidine (DAB) is toxic if swallowed, in contact with skin or if inhaled. It may cause cancer and damage to organs.

BEFORE START INSTRUCTIONS

Cut Formalin-Fixed Paraffin-Embedded (FFPE) tissue specimen at 4 μ m thickness using a water fall microtome.

Deparaffinization

- 1 Dry paraffin sections at room temperature overnight.
- 2 Bake the paraffin sections from 12:00:00 to 24:00:00 at 50 °C .
- 3 Deparaffinization and hydration are done in xylene and graded ethanol to distilled water.

- 4 During hydration step, block endogeneous peroxidase during ⌚ 00:05:00 in H2O2
[M] .3 % volume
in 95% ethanol.

Standard antigen retrieval method

- 5 Heat the slides immersed in retrieval buffer for ⌚ 00:04:00 at 🌡 125 °C in a pressure boiler.
- 6 After completed boiling, leave the slides in the pressure boiler and let them cool down till
🌡 90 °C .
Total time ⌚ 00:45:00

Immunohistochemical staining program, Autostainer 480

- 7 Place slides in the Autostainer 480, all the incubations are done at room temperature.

Equipment

Autostainer 480

NAME

ThermoFisher Scientific

BRAND

cat# A80500007

SKU

<https://www.thermofisher.com/order/catalog/product/A80500008>

LINK

- 8 Rinse slides in wash buffer.

- 9 Incubate slides with Ultra V Block for 00:05:00 .
- 10 Rinse slides in wash buffer go to step #10 once more, total washes 2 .
- 11 Incubate slides with primary antibody diluted in antibody diluent for 00:30:00 .
- 12 Rinse slides in wash buffer go to step #12 two times more, total washes 3 .
- 13 Incubate slides with labeled HRP polymer for 00:30:00 .
- 14 Rinse slides in wash buffer go to step #14 once more, total washes 2 .
- 15 Incubate slides with DAB solution for 00:05:00 .
- 16 Rinse slides in distilled water.

Counterstaining and coverslipping

- 17 Transfer slides in the Autostainer XL.

Equipment

Autostainer XL	NAME
Leica biosystems	BRAND
cat# Leica ST5010 Autostainer XL	SKU
https://www.leicabiosystems.com/histology-equipment/he-stainers-special-stainers-coverslippers/products/leica-st5010/	LINK

- 18 Counterstain slides with hematoxylin for 00:07:50 .

- 19 Rinse slides in lithium carbonate water diluted 1:5 from saturated solution for 00:01:00 .

- 20 Rinse slides in tap water for 00:05:00 .

- 21 Dehydrate slides in graded ethanol and Neo-Clear.

22 Mount coverslip in each of the slide.