MAY 05, 2020

Starting a MinION sequencing run using MinKNOW

Forked from [Starting a MinION sequencing run using MinKNOW](#)

Muhammad Faisal¹, Olin Silander¹, Nikki Freed¹

¹Massey University

Coronavirus Method Development Community

Nikki Freed

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.bfwqjpdw

Protocol Citation: Muhammad Faisal, Olin Silander, Nikki Freed 2020. Starting a MinION sequencing run using MinKNOW. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bfwqjpdw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: In development
We are still developing and optimizing this protocol

Created: May 02, 2020

Last Modified: May 05, 2020

PROTOCOL integer ID:
36528

- 1 If required plug the MinION into the computer and wait for the MinION and flowcell to be detected.

- 2 Choose flow cell 'FLO-MIN106' from the drop-down menu.
- 3 Then select the flowcell so a tick appears.
- 4 Click the 'New Experiment' button in the bottom left of the screen.
- 5 On the New experiment popup screen, select the running parameters for your experiment from the individual tabs:

Experiment: Name the run in the experiment field, leave the sample field blank.

Kit: Selection: Select RBK004

Run Options: Set the run length to 6 hours (you can stop the run once sufficient data has been collected as determined using RAMPART).

Basecalling: Select 'fast basecalling'.

Output: The number of files that MinKNOW will write to a single folder. By default this is set to 4000 but can be reduced to make RAMPART update more frequently.

Click 'Start run'.

- 6 Monitor the progress of the run using the MinKNOW interface.