VERSION 1

AUG 27, 2021

OPEN ACCESS

DOI:

[dx.doi.org/10.17504/protocols.io.buvenw3e](https://doi.org/10.17504/protocols.io.buvenw3e)

Protocol Citation: Peter H L Krijger, Tim A Hoek, Sanne Boersma, Lieke I P M Donders, Maaïke M C Broeders, Mark Pieterse, Pim W Toonen, Ive Logister, Bram M P Verhagen, Marjon J A M Verstegen, Thomas W van Ravesteyn, Rene J T M Roymans, Francesca Mattioli, Jo Vandesompele, Monique Nijhuis, Stefan Meijer, Anton van Weert, Edwin Dekker, Fred J Dom, Rob Ruijtenbeek, Lieven B J van der Velden, Jeroen H B van de Bovenkamp, Martijn Bosch, Wouter de Laat, Marvin E Tanenbaum 2021. Step 1: Swab sample collection.

protocols.io

<https://dx.doi.org/10.17504/protocols.io.buvenw3e>

MANUSCRIPT CITATION:

<https://doi.org/10.1038/s41587-021-01080-6>

🌐 Step 1: Swab sample collection V.1

👤 In 1 collection

Peter H L Krijger¹, Tim A Hoek¹, Sanne Boersma¹, Lieke I P M Donders², Maaïke M C Broeders², Mark Pieterse¹, Pim W Toonen¹, Ive Logister¹, Bram M P Verhagen¹, Marjon J A M Verstegen¹, Thomas W van Ravesteyn¹, Rene J T M Roymans², Francesca Mattioli¹, Jo Vandesompele³, Monique Nijhuis⁴, Stefan Meijer⁵, Anton van Weert⁵, Edwin Dekker⁶, Fred J Dom⁷, Rob Ruijtenbeek⁸, Lieven B J van der Velden², Jeroen H B van de Bovenkamp², Martijn Bosch⁸, Wouter de Laat¹, Marvin E Tanenbaum¹

¹Oncode Institute, Hubrecht Institute-KNAW and University Medical Center Utrecht, Utrecht, the Netherlands.;

²Laboratory of Medical Microbiology, Stichting PAMM, Veldhoven, the Netherlands.;

³InActiv Blue, Beernem, Belgium. Biogazelle, Zwijnaarde, Belgium.;

⁴Dept of Medical Microbiology, University Medical Center, Utrecht, the Netherlands.;

⁵Tecan labwerx, Tecan Trading AG, Männedorf, Switzerland.;

⁶Bodegro B.V., Breda, the Netherlands.;

⁷HiFiBio Therapeutics, Saint-Jacques, France.;

⁸Department of Laboratory Automation, Genmab B.V., Utrecht, the Netherlands.

Peter Krijger

DISCLAIMER

J.V. is co-founder and CSO of InActiv Blue; S.M. is an employee of Tecan Trading AG; L.M.P.D., M.M.C.B., R.J.T.M.R., L.B.J.V. and J.H.B.B. are employees of Stichting PAMM Veldhoven; A.vW. is an employee of Sanquin; E.D. is a founder and employee of Bodegro; and F.J.D. is CEO and co-founder HiFiBiO France.

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

Created: May 10, 2021

Last Modified: Aug 27, 2021

PROTOCOL integer ID:
49798

ABSTRACT

One of the most critical innovations within STRIP is the collection of samples in small tubes (0.8 ml compared with 5–15 ml in standard diagnostics) with virus-inactivating and RNA-stabilizing lysis buffer (InActiv Blue) spiked with a control RNA (PDV). The use of small tubes enables the collection of sample tubes in small 96-tube boxes at sample collection sites. These 96-tube boxes are directly compatible with automated sample processing in the lab without slow and often laborious transfer of samples from collection tubes to plates that are compatible with robotics.

MATERIALS

Sample collection tubes

- 0.8 ml 2D Data-Matrix coded tube with external thread, Micronic, #MP52720.
- Micronic tube rack: Micronic, #MP51207.
- Micronic tube manual recapper: Micronic, #MP35480.
- Micronic tube barcode scanner: Single TUBE HQ SCANNER, HQ healthcare, #268821.
- InActive Blue, virus inactivating and RNA stabilizing transport medium. InActiv Blue, #IB1000.
- Virus transport buffer: InActive Blue mixed with a spike-in control RNA (We use home-made Phocine Distemper Virus, PDV).

0.8 ml Micronic tubes are filled by Micronic with 0.4 ml virus transport buffer and can be stored at RT for at least 8 months. Before shipment to the collection sites the tubes are centrifuged (1 min 500G).

SAFETY WARNINGS

- ❗ InActive Blue contains Guanidine thiocyanate. Guanidine thiocyanate is harmful if inhaled, swallowed or if it comes into contact with skin and causes severe skin burns and eye damage. In addition, never clean up InActive Blue with bleach as Guanidine thiocyanate reacts with bleach to produce cyanide gas.

BEFORE START INSTRUCTIONS

Collection personnel should be trained in the proper collection procedure.

Sample collection

- 1** Person 1 scans the 2D barcode at the bottom of a 0.8 ml Micronic tube prefilled with 0.4 ml virus transport buffer and places it in a Micronic tube rack (tube rack A).
- 2** Person 1 removes cap from the Micronic tube using the manual recapping device (cap stays on the recapping device).
- 3** After instructing the patient to blow their nose, the naso- oropharyngeal swab specimen is collected by person 2 using 1 flocked swab using standard procedures. Importantly: Make sure no mucus is attached. Repeat swab collection when needed.
- 4** Person 2 places the swap in the tube placed in tube rack A, rotates the swab 10 times between thumb and index finger, and trashes the swab.
- 5** Person 1 close the tube with the cap and transfers the tube to a new tube collection rack (tube rack B).
- 6** Store at room temperature until shipment to the RNA extraction site. Samples are processed as soon as possible after arrival at the lab. Samples are regularly processed within 24h but can be stored up to 72h if required.