MAR 02, 2020

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.974h9qw

Protocol Citation: Omneya Ahmed Osman, Alexander Eiler, Mats Töpel, Tomas Iarsson 2020. The detection of freshwater pearl mussel . **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.974h9qw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Dec 06, 2019

Last Modified: Mar 02, 2020

PROTOCOL integer ID:
 30684

Keywords: freshwater pearl mussel, qPCR

🌐 The detection of freshwater pearl mussel

Omneya Ahmed Osman¹, Alexander Eiler¹, Mats Töpel¹, Tomas Iarsson¹

¹eDNA solutions

eDNASolutions

Tech. support phone: **+0700264843** email: **omneya@ednasolutions.se**

Omneya Ahmed Osman

ABSTRACT

qPCR assay to detect freshwater pearl mussels *Margaritifera margaritifera* in order to monitor populations in their natural environment.

ATTACHMENTS

[qPCR_Freshwater PearlMussel.xlsx](#)

GUIDELINES

Laboratory work space and equipment were sterilized by UV-light and DNase solution and 70% ethanol. Filter pipet tips were used in all steps of the laboratory work.

Negative controls of DNase/RNase free water were used in each qPCR assay.

MATERIALS

MATERIALS

UltraPure™ DEPC-treated Water **Thermo Fisher Catalog #10813999**

Qiagen DNeasy power water sterivex kit **Contributed by users Catalog #14600-50-NF**

STEP MATERIALS

SsoAdvanced Universal SYBR® Green Supermix **Bio-Rad Laboratories Catalog #172-5270**

SsoAdvanced Universal SYBR® Green Supermix **Bio-Rad Laboratories Catalog #172-5270**

SAFETY WARNINGS

- Handling high concentration of positive controls was performed in a post-PCR room which is physically separated from the pre-PCR room to avoid contamination.
Always add your samples first and seal them before adding the serial dilutions of positive control (standard) at the end.

- 1 DNA extraction was performed using Qiagen DNeasy power water sterivex kit. The quality of the extracted DNA was estimated using Nanodrop.

Qiagen DNeasy power water sterivex protocol can be found in this link:

<https://www.qiagen.com/se/resources/resourcedetail?id=c5fe7d5f-070a-4ebe-ac04-4bbf05a13e91&lang=en>

Primers

- 2 Stoeckle et al., 2016 designed primers targeting 16S rRNA gene of *Margaritifera margaritifera*. The sequence of the primer set is:
MarMa_16S2.1: 5'- GCAACACGGAAAACCCC TG -3'
MarMa_16S1.2: 5'- GGCT GCGCTCATGTGAATTA -3'.

- 3 The concentration of freshwater pearl mussel standard was measured by PicoGreen assay (<http://tools.thermofisher.com/content/sfs/manuals/PicoGreen-dsDNA-protocol.pdf>) before

preparing standard serial dilutions.

Serial dilutions, 10, 1, 0.1, 0.01 and 0.001 ng, were prepared to generate standard curve in each qPCR run.

Unio pictorum or Äkta målar-mussla was used to test the specificity of qPCR detection.

Plate setup

- 4 Setup and design your qPCR plate in the template of qPCR machine before preparing your qPCR run.

At least 3 or 4 replica of standard dilutions should be included in each run.

Three replica of each sample was performed.

PCR reaction

- 5 A master mix was prepared and calculated according to the number of standard dilutions, samples and negative controls to run.

The PCR mixture per sample was as follows:

Reagent	working sol	final conc	μl
SsoAdvanced Universal SYBR® Green Supermix	2x	1x	10
MarMa_16SF primer	10 uM	0,25 uM	0,5
MarMa_16SR primer	10 uM	0,25 uM	0,5
DEPC-water			6
DNA template			3
Σ			20

Sample or qPCR plate should be carefully sealed with no marking on the tube to facilitate plate reader step.

PCR program

6

Amplification step	Time	Temp	

Enzyme activation	15 sec	98 °C	
Denaturation	10 sec	98 °C	40 cycles
Annealing	10 sec	60 °C	
Extension	20 sec	72 °C	
Melt curve analysis		63 to 99°C	

7

QPCR was performed in BioRad qPCR machine CFX96.

Analysis of the results was done by CFX maestro software

<https://www.bio-rad.com/en-se/product/cfx-maestro-software-for-cfx-real-time-pcr-instruments?ID=OKZP7E15>

QPCR efficiency, limit of detection and quantification range of standard were checked for each run.

Results can be evaluated through data analysis windows where there is a quantification, standard melt curve and RFU spreadsheet.