NOV 07, 2017

OPEN ACCESS

DOI:
[dx.doi.org/10.17504/protocols.io.ki7cuhn](https://doi.org/10.17504/protocols.io.ki7cuhn)

External link:
<https://doi.org/10.1371/journal.pone.0193657>

Protocol Citation: Sven Van Poucke, Dana Huskens, Kurt Van der Speeten, Mark Roest, Bart Lauwereins, Ming-Hua Zheng, Seppe Dehaene, Joris Penders, Abraham Marcus, Marcus Lancé 2017. Thrombin generation assay (CAT). **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.ki7cuhn>

MANUSCRIPT CITATION:
 Poucke SV, Huskens D, Speeten KVd, Roest M, Lauwereins B, Zheng M, Dehaene S, Penders J, Marcus A, Lancé M (2018) Thrombin generation and platelet activation in cytoreductive surgery combined with hyperthermic intraperitoneal chemotherapy - A prospective cohort study. PLoS ONE 13(6): e0193657. doi: [10.1371/journal.pone.0193657](https://doi.org/10.1371/journal.pone.0193657)

Thrombin generation assay (CAT)

PLOS One

Sven Van Poucke¹, Dana Huskens², Kurt Van der Speeten³, Mark Roest², Bart Lauwereins¹, Ming-Hua Zheng⁴, Seppe Dehaene¹, Joris Penders⁵, Abraham Marcus⁶, Marcus Lancé⁶

¹Department of Anesthesiology, Intensive Care, Emergency Medicine and Pain Therapy, ZOL, Genk, Belgium.;

²Synapse Research Institute, Maastricht, The Netherlands.;

³Department of Surgery, ZOL, Genk, Belgium;

⁴Department of Hepatology, the First Affiliated Hospital of Wenzhou Medical University, Wenzhou 325000, China.;

⁵Central Diagnostic Laboratory, ZOL, Genk, Belgium;

⁶Department of Anesthesiology, ICU and Perioperative Medicine, HMC, Doha, Qatar

HIPACAT

Sven Van Poucke

Department of Anesthesiology, Intensive Care, Emergency Medi...

ABSTRACT

TG in plasma was measured with the calibrated automated thrombogram (CAT) assay as developed by Hemker and co-workers [18-20]. Briefly, 80 µl platelet poor plasma (PPP) was mixed with 20 µl of a mixture containing tissue factor (Dade-Behring) at a final concentration of 1 pM and phospholipid vesicles (f.c. 4 µM 20 mol% phosphatidylserine, 60 mol% phosphatidylcholine and 20 mol% phosphatidylethanolamine, Avanti). To calibrator wells, 20 µl of calibrator (α2macroglobulin-thrombin complex, [19]) was added instead of TF and PL. After 10 minutes of incubation at 37°C, thrombin generation was initiated by the addition of 20 µl of the thrombin specific substrate, Z- Gly-Gly-Arg-7-amino-4-methylcoumarin (f.c. 416 µM, Bachem) and CaCl₂ (f.c. 16.7 mM). Fluorescence was measured with a Fluoroscan Ascent reader (Thermo Labsystems) and data were analyzed with dedicated software (Thrombinoscope, Stago) [20]. Thrombin generation was expressed based on endogenous thrombin potential (ETP); lagtime (LT); thrombin peak (TP), time-to-thrombin peak (TTP).

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

Created: Nov 01, 2017

Last Modified: Mar 03, 2018

PROTOCOL integer ID:
8511

Keywords: Thrombin