

Nov 07, 2017

Thrombin generation assay (CAT)

 [PLOS One](#)

DOI

dx.doi.org/10.17504/protocols.io.ki7cuhn

Sven Van Poucke¹, Dana Huskens², Kurt Van der Speeten³, Mark Roest², Bart Lauwereins¹, Ming-Hua Zheng⁴, Seppe Dehaene¹, Joris Penders⁵, Abraham Marcus⁶, Marcus Lancé⁶

¹Department of Anesthesiology, Intensive Care, Emergency Medicine and Pain Therapy, ZOL, Genk, Belgium.;

²Synapse Research Institute, Maastricht, The Netherlands.; ³Department of Surgery, ZOL, Genk, Belgium;

⁴Department of Hepatology, the First Affiliated Hospital of Wenzhou Medical University, Wenzhou 325000, China.;

⁵Central Diagnostic Laboratory, ZOL, Genk, Belgium;

⁶Department of Anesthesiology, ICU and Perioperative Medicine, HMC, Doha, Qatar

HIPACAT



Sven Van Poucke

Department of Anesthesiology, Intensive Care, Emergency Medi...

OPEN  ACCESS



DOI: dx.doi.org/10.17504/protocols.io.ki7cuhn

External link: <https://doi.org/10.1371/journal.pone.0193657>

Protocol Citation: Sven Van Poucke, Dana Huskens, Kurt Van der Speeten, Mark Roest, Bart Lauwereins, Ming-Hua Zheng, Seppe Dehaene, Joris Penders, Abraham Marcus, Marcus Lancé 2017. Thrombin generation assay (CAT). **protocols.io** <https://dx.doi.org/10.17504/protocols.io.ki7cuhn>

Manuscript citation:

Poucke SV, Huskens D, Speeten KVd, Roest M, Lauwereins B, Zheng M, Dehaene S, Penders J, Marcus A, Lancé M (2018) Thrombin generation and platelet activation in cytoreductive surgery combined with hyperthermic intraperitoneal chemotherapy - A prospective cohort study. PLoS ONE 13(6): e0193657. doi: [10.1371/journal.pone.0193657](https://doi.org/10.1371/journal.pone.0193657)

License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited



Protocol status: Working

Created: November 01, 2017

Last Modified: March 03, 2018

Protocol Integer ID: 8511

Keywords: Thrombin

Abstract

TG in plasma was measured with the calibrated automated thrombogram (CAT) assay as developed by Hemker and co-workers [18-20]. Briefly, 80 μ l platelet poor plasma (PPP) was mixed with 20 μ l of a mixture containing tissue factor (Dade-Behring) at a final concentration of 1 pM and phospholipid vesicles (f.c. 4 μ M 20 mol% phosphatidylserine, 60 mol% phosphatidylcholine and 20 mol% phosphatidyl-ethanolamine, Avanti). To calibrator wells, 20 μ l of calibrator (α 2macroglobulin- thrombin complex, [19]) was added instead of TF and PL. After 10 minutes of incubation at 37°C, thrombin generation was initiated by the addition of 20 μ l of the thrombin specific substrate, Z- Gly-Gly-Arg-7-amino-4-methylcoumarin (f.c. 416 μ M, Bachem) and CaCl₂ (f.c. 16.7 mM). Fluorescence was measured with a Fluoroscan Ascent reader (Thermo Labsystems) and data were analyzed with dedicated software (Thrombinoscope, Stago) [20]. Thrombin generation was expressed based on endogenous thrombin potential (ETP); lagtime (LT); thrombin peak (TP), time-to-thrombin peak (TTP).

