FEB 25, 2021

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.bspyndpw

Protocol Citation: John Herndon, Madelyn Carmody, Stephen Oh, William Gillanders Washington University St Louis 2021. Tissue Staining for Imaging Mass Cytometry. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.bspyndpw>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Feb 23, 2021

Last Modified: Feb 25, 2021

PROTOCOL integer ID:
 47576

Tissue Staining for Imaging Mass Cytometry

John Herndon¹, Madelyn Carmody², Stephen Oh²,
 William Gillanders Washington University St Louis²

¹Washington University in Saint Louis, Department of Surgery;

²Washington University, Saint Louis

NCIHTAN

John Herndon

Washington University in Saint Louis, Department of Surgery

ABSTRACT

This protocol describes the preparation and staining of human FFPE material for multiplex visualization using Imaging Mass Cytometry (IMC). This protocol uses basic standard immunohistochemical staining techniques. The tissue is incubated with antibodies that have specific affinity for different cells and tissues in the context of a formalin fixed thin section of human tumor material. Antibodies are labeled with heavy metals instead of the typical fluorochromes, and are visualized using the Hyperion Imaging System manufactured by the Fluidigm corporation.

GUIDELINES

This technique uses specialized metal tagged antibodies that can be purchased or conjugated by the end user. Dilutions must be empirically determined and validated before using in this protocol. Antibodies are prepared as a cocktail and incubated on the slides all at once.

All containers for reagents, etc. must be in plastic. Glass hardware must not be used, i.e. graduated cylinders, pipettes, bottles, etc. Glass contains lead at low levels and can interfere with heavy metal tagged antibodies.

Slides used must also be kept desiccated and should ideally be freshly cut, do not use any slides that have been cut from blocks for greater than two weeks, if possible. Oxidation and moisture can interfere with the visualization of the heavy metal tagged antibodies.

MATERIALS

Xylene, histology grade

Ethanol 100% and diluted in diH₂O

Tris EDTA pH 9.0 with 10% Glycerol (can be purchased pre made and add glycerol)

Glycerol

PAP pen

Phosphate Buffered Saline (PBS)

Triton X-100

Bovine Serum Albumin (BSA)

DNA intercalator-Ir, **1mM 500 micromolar (μM)** obtain from Fluidigm, catalog # 201192B

Ruthenium Red, Sigma 00541-1G

SAFETY WARNINGS

Xylene and ethanol are volatile chemicals with an inhalation risk. All steps performed with Xylene and Ethanol should be in a certified chemical fume hood.

BEFORE START INSTRUCTIONS

Be sure to use clean fresh reagents.

1

Place positively charged slides with **→|← 5 μm** thick tissue section in a **🔥 60 °C** oven for **🕒 02:00:00**

2h

2

Place slides in fresh xylene for **🕒 00:20:00**

20m

3

Turn on Water bath for Antigen Retrieval, **🔥 96.0 °C**

4

Prepare Antigen retrieval solution of Tris EDTA **📏 pH 9.0** with 10% Glycerol

5 Place container for slides with above antigen retrieval solution in to water bath to pre-warm

6 **Rehydrate**

Place slides in to 100% Ethanol, 00:05:00

5m

7 Places slides in to 95% Ethanol, 00:05:00

5m

8 Placed slides in to 80% Ethanol, 00:05:00

5m

9 Place slides in to 70% Ethanol, 00:05:00

5m

10 Place slides in to distilled water for 00:05:00

5m

* Note slides may be kept for a longer period at this step, make sure the water bath is the correct temperature before proceeding to the next step.

11 **Antigen Retrieval**

30m

Incubate slides for 00:30:00 in antigen retrieval solution from step 4

Make sure the 96.0 °C is maintained consistently throughout the 30 minute incubation.

12 After incubation, cool slides to 70 °C on the bench top (about 10 minutes)

- 13 Wash** 10m
Place slides in distilled water for 00:10:00 with gentle agitation
- 14** repeat Step 13, 00:10:00 10m
- 15** Place slides in PBS for 00:10:00 with gentle agitation 10m
- 16** repeat Step 15 00:10:00 10m
- 17 Block Nonspecific Staining**
Encircle each sample with a PAP pen
- 18** Blocking solution is 3% BSA in PBS, add just enough to cover the sample and incubate for 00:45:00 at Room temperature in a hydration chamber. 45m
- 19 Antibody Staining**
Pipette 50-100 μL pre-prepared antibody cocktail in 0.5% BSA and incubate Overnight at 4.0 °C in hydration chamber
- 20 Wash** 8m
Wash slides with PBS with 0.2% Triton X-100 for 00:08:00 with gentle agitation

- 21 Repeat Step 20, 00:08:00 8m
- 22 Wash slides in PBS for 00:08:00 with gentle agitation 8m
- 23 repeat step 22, 00:08:00 8m
- 24 **DNA Intercalator-Ir** 30m
Prepare Iridium intercalator by diluting 1 to 1000 in PBS (this dilution may have to determined empirically)
Pipette just enough iridium intercalator to cover sample and incubate for 00:30:00 at
Room temperature
- 25 Remove iridium intercalator by gently tapping edge of slide to a paper towel
- 26 Wash slides with PBS for 00:08:00 with gentle agitation 8m
- 27 **Ruthenium Counter Stain** 10m
Prepare a stock solution of Ruthenium Red by mixing 0.0025 grams in PBS, vortex well to mix
Dilute Ruthenium to working concentration of 0.0005%
Counterstain tissue by pipetting just enough of the 0.0005% ruthenium red in PBS for
00:10:00 (duration of ruthenium red may need to be determined empirically)
- 28 Remove ruthenium red solution by gently tapping edge of slide to paper towel

- 29 Wash** 5m
Wash slides with distilled water for 00:05:00 with gentle agitation
- 30** Air dry slides at room temperature for 00:20:00 20m
- 31** Slides are now ready for Hyperion Imaging.