



Sep 10, 2024

Version 2

# Tomogram Reconstruction and Gold Fiducial Removal with WarpTools, Etomo, and Fidder V.2

DOI

<https://dx.doi.org/10.17504/protocols.io.6qpvr8qbblmk/v2>

Connor Garrels<sup>1</sup>, Benjamin Barad<sup>1</sup>, Steve Reichow<sup>1</sup>

<sup>1</sup>Oregon Health & Science University



Connor Garrels

Oregon Health & Science University

## Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



**DOI:** <https://dx.doi.org/10.17504/protocols.io.6qpvr8qbblmk/v2>

**Protocol Citation:** Connor Garrels, Benjamin Barad, Steve Reichow 2024. Tomogram Reconstruction and Gold Fiducial Removal with WarpTools, Etomo, and Fidder. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.6qpvr8qbblmk/v2> Version created by **Connor Garrels**

**License:** This is an open access protocol distributed under the terms of the **[Creative Commons Attribution License](#)**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

**Protocol status:** Working

**We use this protocol and it's working**

**Created:** September 06, 2024

**Last Modified:** September 10, 2024

**Protocol Integer ID:** 107070

**Keywords:** cryoET, tomography, image processing, python, gold fiducial removal with warptool, fidder while gold nanoparticle fiducial, gold fiducial removal, electron tomography data, gold nanoparticle fiducial, tomogram reconstruction, fidder into tomogram reconstruction, fiducial removal, quality fiducial removal, series alignment in cryo, warptool, using warp, cryo, using fidder

## Abstract

While gold nanoparticle fiducials greatly improve tilt-series alignment in cryo-electron tomography data, they can cause artifacts during later image processing steps. We present a workflow for high-quality fiducial removal using fidder, an open-source python package. This protocol demonstrates how to flexibly incorporate fidder into tomogram reconstruction using Warp and Etomo.

## Overview

- 1 Gold nanoparticle fiducials in cryoET can greatly improve tilt-series alignment and tomogram quality. They can also introduce detrimental artifacts in downstream processing steps such as denoising or membrane segmentation. While there are methods to remove gold beads from micrographs, these methods are not always high quality or easily integrated into existing workflows.

Here, we describe a workflow using *fitter* as part of tomogram reconstruction with WarpTools and Etomo, incorporating a step for gold fiducial removal with *Fidder*. The resulting tomograms are high quality because of the ability to use gold during alignment, but with no visible gold artifacts remaining.

## Install Software

- 2 Make sure you have these packages installed.
  - [IMOD](#)
  - [WarpTools](#)
  - [Fidder](#)

## WarpTools Preprocessing

- 3 We will be using WarpTools for the bulk of our processing. This section follows [this warptools tutorial](#).

Values shown are data-specific (i.e. exposure, pixel size, paths), so please be careful when copying these commands. These are just example values!

### 3.1 Create Settings Files

frameseries

```
WarpTools create_settings --folder_data /where/your/movies/are --
folder_processing frameseries --output frameseries.settings --
extension *.tif --angpix 0.8125 --bin 1 --gain_path
/where/your/gain/is --exposure 3
```

tiltseries

```
WarpTools create_settings --output tiltseries.settings --  
folder_processing tiltseries --folder_data tomostar --extension  
*.tomostar --angpix 0.8125 --bin 1 --gain_path /where/your/gain/is  
--exposure 3 --tomo_dimensions 9600x12000x2500
```

If your data was collected at super-resolution like ours was, --bin 1 is very important!

### 3.2 Motion Correction and CTF Estimation

Motion correct and CTF estimate your data and output whole frame averages as well as even and odd frame-split micrographs for later denoising.

```
WarpTools fs_motion_and_ctf --settings frameseries.settings --  
m_grid 8x6x10 --c_grid 2x2x1 --c_range_max 7 --c_defocus_max 8 --  
c_use_sum --out_averages --out_average_halves
```

### 3.3 Import tiltseries metadata

```
WarpTools ts_import --mdocs /where/your/mdocs/are --frameseries  
frameseries --tilt_exposure 3 --min_intensity 0.3 --dont_invert --  
output tomostar
```

You may need to edit your mdocs to reflect where your motion corrected micrographs are. If so, try this script.

```
#!/usr/bin/env python3

import os

mdoc_folder = '/where/your/mdocs/are'
new_subframe_path = 'where/your/micrographs/are'

def update_subframe_path(mdoc_path, new_path):
    with open(mdoc_path, 'r') as f:
        lines = f.readlines()

    for i, line in enumerate(lines):
        if line.startswith('SubFramePath'):
            existing_path = line.split('=')
            [-1].strip().replace('\\', '/')
            _, existing_filename = os.path.split(existing_path)

            new_subframe_path = f'SubFramePath = {os.path.join(new_path, existing_filename)}\n'
            lines[i] = new_subframe_path

    with open(mdoc_path, 'w') as f:
        f.writelines(lines)

def batch_update_paths(folder_path, new_path):
    for filename in os.listdir(folder_path):
        if filename.endswith('.mdoc'):
            mdoc_path = os.path.join(folder_path, filename)
            update_subframe_path(mdoc_path, new_path)
            print(f'SubFramePath updated for {filename}')

batch_update_paths(mdoc_folder, new_subframe_path)
```

### 3.4 Make Stacks

```
WarpTools ts_stack --settings tiltseries.settings --angpix 6.5
```

Notice we chose to make our stacks at 6.5 Å (bin 4 from physical pixel size). Whatever size you choose, remember it, as it is the resolution you will use for alignment and reconstruction later.



## Etomo Tilt Series Alignment

- 4 We will use batch processing in etomo for aligning our newly created stacks. Components of this section follows [this warp tutorial](#).

### 4.1 Make etomo directory

Create the directory 'etomo' in your /tiltseries directory.

### 4.2 etomo

Open etomo in your /tilseries/etomo directory and choose the 'Batch Processing' button.

Set up your etomo session thusly:

Batch Setup Tab

- Select 'Move all stacks to dataset directory under:' and enter the path to your /tiltseries/etomo directory

Stacks Tab

- Add the stacks made in the last Warp step. The stacks will be .st files.

Dataset Values (Basic Tab)

- Set fiducial bead size in nm (my BSA-coated 5nm gold beads are closer to 8nm)
- 'Coarse aligned stack binning - single frame' should be 1 (no binning)
- 'Aligned stack binning' should be 1 (no binning)
- Select 'R-weighted back projection' and deselect 'SIRT-like filter'

Dataset Values (Advanced Tab)

- Set 'Tilt Axis Rotation' (mine was 85)
- 'Use existing .rawtilt file for A/only axis:' should be set to Yes. (**IMPORTANT**)
- 'Type of magnification solution' should be 0: *fixed*
- 'Type of rotation solution' should be 5: *grouped*
- 'Type of beam tilt solution' should be -1: *solve one*

Run Tab

- Create the file /usr/local/ImodCalib/cpu.adoc [per these instructions](#).

 cpu.adoc 64B

'number' is the number of logical cpu cores on your machine. 'gpu.device' is the list of GPUs on your machine (note the list starts at 1, unlike CUDA device nomenclature).

- With the .adoc created, select your desired # CPUs value (I chose 56).

- Select 'Parallel GPUs' and set your # *GPUs* value (I chose 2)
- Select 'Run multiple batch jobs in parallel' if you dare. I was able to run 7 jobs with a maximum of 2 GPUs per job.
- Under 'Subset of Steps to Run', check *Stop after* and select *Fine alignment*.  
**(IMPORTANT)**
- Click *Run* and rock and roll

This will output two alignments files for each tiltseries - .xf and .tlt.

## WarpTools Importing Alignments

- 5 Head back to warp to import the etomo alignments and get ready for reconstruction. Still following [this guide](#).

### 5.1 Import alignments

Copy the .xf and .tlt files from their etomo subdirectories and move to their own folder at tiltseries/imod\_xf\_tlt.

Then run

```
WarpTools ts_import_alignments --settings tiltseries.settings --alignments tiltseries/imod_xf_tlt --alignment_angpix 6.5
```

Remember we made our stacks at 6.5Å in step 2.4 and did no binning in etomo, so the alignments were all calculated at 6.5Å.

### 5.2 Check Handedness

Back to [this tutorial](#).

```
WarpTools ts_defocus_hand --settings tiltseries.settings --check
```

Follow the instructions in the output of this job to change your handedness if necessary.

### 5.3 CTF Estimation

```
WarpTools ts_ctf --settings tiltseries.settings --range_high 7 --defocus_max 8
```



## Fidder Gold Removal

- 6 **Fidder** - involves two steps; creating masks of fiducials and erasing them. We have provided an example python script to complete both steps on your full and half micrographs so long as they are within the warp file structure. There is also the option to use the command-line for fidder.
- 6.1 You should only have to edit the parent\_dir variable.

```
#!/usr/bin/env python3

import os
from multiprocessing import Pool
import mrcfile
import torch
from fidder.predict import predict_fiducial_mask
from fidder.erase import erase_masked_region

parent_dir = '/path/to/you/warp/frameseries/'

def check_directories(parent_dir):
    for subdir in subdirs:
        dir_path = os.path.join(parent_dir, subdir)
        if os.path.exists(dir_path):
            print(dir_path + ' found.')
        if not os.path.exists(dir_path):
            os.makedirs(dir_path)
            print(dir_path + ' created.')

    os.makedirs(os.path.join(parent_dir, 'average_erased'),
exist_ok=True)
    os.makedirs(os.path.join(parent_dir, 'average',
'even_erased'), exist_ok=True)
    os.makedirs(os.path.join(parent_dir, 'average', 'odd_erased'),
exist_ok=True)

def make_mask(filename, parent_dir):
    """ Apply fidder's predict_fiducial_mask function to a single
micrograph and save the resultant mask as a new mrc file.

    Args:
        filename (str) : The name of the micrograph to process.
        parent_dir (str) : The directory containing micrographs in
/average and frame-split halves in
/average/even, and /average/odd subdirectories.
    """
    mic_path = os.path.join(parent_dir, 'average', filename)
    mask_path = os.path.join(parent_dir, 'average', 'mask',
filename)

    image = torch.tensor(mrcfile.read(mic_path))
```

```
mask, probabilities = predict_fiducial_mask(
    image, pixel_spacing=1.35, probability_threshold=0.7
)

mask_uint8 = mask.to(torch.uint8)

os.makedirs(os.path.dirname(mask_path), exist_ok=True)

with mrcfile.new(mask_path, overwrite=True) as mrc:
    mrc.set_data(mask_uint8.numpy())

print(mask_path + '/' + filename + ' mask made.')

def erase_gold(filename, parent_dir, subdir):
    """Apply fidder's erase_masked_region function to a single
    frame and save the result as a new mrc file.

    Args:
        filename (str): The name of the micrograph to process.
        parent_dir (str) : The directory containing micrographs in
        /average and frame-split halves in
        /average/even, and /average/odd subdirectories.
    """
    mask_path = os.path.join(parent_dir, 'average', 'mask',
                             filename)

    mic_path = os.path.join(parent_dir, subdir, filename)

    if subdir == 'average':
        output_subdir = os.path.join(parent_dir, 'average_erased')
    else:
        output_subdir = os.path.join(parent_dir, subdir +
        '_erased')

    os.makedirs(output_subdir, exist_ok=True)

    image = torch.tensor(mrcfile.read(mic_path))
    mask = torch.tensor(mrcfile.read(mask_path))

    erased_image = erase_masked_region(image=image, mask=mask)

    mrc_output_path = os.path.join(output_subdir, filename)
    with mrcfile.new(mrc_output_path, overwrite=True) as mrc:
        mrc.set_data(erased_image.numpy())
```

```
        print(output_subdir + '/' + filename + ' completed')

def process_gold(subdir, parent_dir):
    filenames = [i for i in os.listdir(os.path.join(parent_dir,
subdir)) if i.endswith('.mrc')]
    with Pool(48) as p:
        p.starmap(erase_gold, [(filename, parent_dir, subdir) for
filename in filenames])
    print('##### all gold erased for '
+ subdir + ' #####')

subdirs = ['average', 'average/even', 'average/odd']

check_directories(parent_dir)

filenames = [i for i in os.listdir(os.path.join(parent_dir,
'average')) if i.endswith('.mrc')]
for filename in filenames:
    mask_path = os.path.join(parent_dir, 'average', 'mask',
filename)
    if not os.path.exists(mask_path):
        make_mask(filename, parent_dir)
    else:
        print(f'{filename} already exists')
print('##### mask processing complete
#####')

for subdir in subdirs:
    process_gold(subdir, parent_dir)
```

Depending on your computational resources, you may want to change the Pool() variable. erase\_gold runs well with Pool(48) on our 64-thread machine.

Rename your existing (gold-containing) micrograph directories so they are not recognized by warp (e.g. /averages becomes /averages\_gold etc.)

Move your newly gold-erased micrographs to the directories you just renamed (e.g. /frameseries/average, /frameseries/average/even, frameseries/average/odd)

## WarpTools Tomogram Reconstruction



- 7 To reconstruct your beautiful gold-erased tomograms:

```
WarpTools ts_reconstruct --settings tiltseries.settings --  
dont_invert --halfmap_frames --angpix 10
```

## Protocol references

Tegunov, D., Cramer, P. Real-time cryo-electron microscopy data preprocessing with Warp. *Nat Methods* **16**, 1146–1152 (2019). <https://doi.org/10.1038/s41592-019-0580-y>

Mastronarde DN, Held SR. Automated tilt series alignment and tomographic reconstruction in IMOD. *J Struct Biol.* 2017 Feb;197(2):102–113. doi: 10.1016/j.jsb.2016.07.011. Epub 2016 Jul 19. PMID: 27444392; PMCID: PMC5247408.

Kremer JR, Mastronarde DN, McIntosh JR. Computer visualization of three-dimensional image data using IMOD. *J Struct Biol.* 1996 Jan-Feb;116(1):71–6. doi: 10.1006/jsbi.1996.0013. PMID: 8742726.