VERSION 4

SEP 09, 2021

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.bujgnujw

Protocol Citation: Adélaïde Roguet 2021. Total nucleic acid extraction - Maxwell(R) HT Environmental TNA Kit, custom (Promega). **protocols.io** <https://dx.doi.org/10.17504/protocols.io.bujgnujw> Version created by **McLellan Lab**

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: In development
We are still developing and optimizing this protocol

Created: Apr 27, 2021

Last Modified: Sep 09, 2021

PROTOCOL integer ID:
49480

🌐 Total nucleic acid extraction - Maxwell(R) HT Environmental TNA Kit, custom (Promega) V.4

Adélaïde Roguet¹

¹UWM

mcLellan lab

McLellan Lab

DISCLAIMER

DISCLAIMER – FOR INFORMATIONAL PURPOSES ONLY; USE AT YOUR OWN RISK

The protocol content here is for informational purposes only and does not constitute legal, medical, clinical, or safety advice, or otherwise; content added to protocols.io is not peer reviewed and may not have undergone a formal approval of any kind. Information presented in this protocol should not substitute for independent professional judgment, advice, diagnosis, or treatment. Any action you take or refrain from taking using or relying upon the information presented here is strictly at your own risk. You agree that neither the Company nor any of the authors, contributors, administrators, or anyone else associated with protocols.io, can be held responsible for your use of the information contained in or linked to this protocol or any of our Sites/Apps and Services.

ABSTRACT

Total nucleic acid extraction from wastewater using Maxwell(R) HT Environmental TNA Kit, custom (Promega)

GUIDELINES

When work is completed, remove equipment and supplies from the cabinet. Wipe the work area with 10% bleach, let stand for 10 min, rinse with water, then with 70% ethanol, and finally with RNAase AWAY.

Keywords: wastewater, SARS-CoV-2, total nucleic extraction, RNA, DNA, extraction, purification, Promega, KingFisher

MATERIALS

MATERIALS

- Ethanol USP/ACS or molecular biology grade (100%)
- Molecular biology grade water
- Isopropanol** molecular biology grade (100%)
- 6x KingFisher 96-well plates** (Cat. no.: 95040460)
- 1x KingFisher 96-tip comb well plate**(Cat. no.: 97002534)
- Screw cap** microcentrifuge tubes

Equipment	
Mini-Beadbeater-16	NAME
high-energy cell disrupter	TYPE
BioSpec	BRAND
607	SKU
https://biospec.com/product/mini-beadbeater-16	LINK
1 speed	SPECIFICATIONS

Equipment	
Kingfisher Flex	NAME
Automated Extraction System	TYPE
ThermoFisher	BRAND
5400630	SKU

Promega_Maxwell_HT_RNA_Wastewater_V1.bdz

BEFORE START INSTRUCTIONS

1. Clean the working area and all equipment: wipe down with 10% bleach and let dry. Wipe down with 70% ethanol and let dry. Then, wipe down using RNase AWAY and let dry.
2. Prepare the 50% ethanol solution (it must be fresh!)
3. Prepare the 6 purification plates:
 - **Wash 1 plate:** Add 100 µl of 50% ethanol and 900 µl of wash buffer (WBA) to each well required for purification.
 - **Wash 2 plate** (same as plate Wash 1): Add 100 µl of 50% ethanol and 900 µl of wash buffer (WBA) to each well required for purification.
 - **Ethanol Wash plate:** Add 450 µl of 50% ethanol to each well required for purification.
 - **Elution plate:** Add 100 µl of 25 mM Tris-HCl (pH 8.0) to each well required for purification.
 - **Tip plate:** Place KingFisher 96-tip comb into an empty KingFisher 96-well plate. *While opening the 96-tip comb plate, pay attention to not touch the tips.*
 - **Lysis and Bind plate:** Add 35 µL of Resin to each well required for purification (*vortex the bottle at max speed before use*). Add 50 µl of Alkaline Protease Solution custom (APA) to each well required for purification. Add 250 µl of cell lysis solution (CLD) to each well required for purification. Add 400 µl of Isopropanol (100%) to each well required for purification.

Total nucleic acid extraction

2h

- 1 For **HA filter** extraction, let the sample thaw on ice and go to **step 2**.

5m

For **BCoV/BRSV** extraction (in duplicate), add 5 µL of BCoV/BRSV solution to the 2-mL tube containing 250 µL CTAB. Vortex for 15 seconds (speed 7 out of 10) and flash freeze the tube. Go to step 4.

For **Direct extraction**, add 150 µL of wastewater to the 2-mL tube containing 100 µL CTAB. Vortex for 15 seconds (speed 7 out of 10) and flash freeze the tube. Go to step 4.

2

For **HA filter** extraction, place the 2-mL tubes in the bead beater.

Equipment

Mini-Beadbeater-16

NAME

high-energy cell disrupter

TYPE

BioSpec

BRAND

607

SKU

<https://biospec.com/product/mini-beadbeater-16>

LINK

1 speed

SPECIFICATIONS

2.1 Bead beat for 00:02:30 2m 30s

Safety information

Start the bead beating when the beads start to be loose in the tubes.

2.2 Cooldown the samples on ice for 00:05:00 5m

2.3 Repeat Steps 9.1 and 9.2 once 7m 30s

3 Centrifuge at maximum speed for 1 min at room temperature. 1m
150000 rpm, Room temperature, 00:01:00

4 For **HA filter** extraction, transfer 125-250 μ L of supernatant to the **Lysis/Bind plate**. 10m
For **B CoV/BRSV/Direct extraction**, transfer all supernatant to the **Lysis/Bind plate**.

Note

The default volume transferred is 250 µL. However, for WWTPs with "dirty" influents, we only transfer 125 µL.

- 5 Start the protocol Promega_Maxwell_HT_RNA_Wastewater_V1.bdz on the KingFisher Flex

1h 14m

🕒 01:14:00

Equipment

Kingfisher Flex	NAME
Automated Extraction System	TYPE
ThermoFisher	BRAND
5400630	SKU

- 6 Transfer the purified sample from the **Elution plate** to the **microcentrifuge tubes**.

10m

Note

The DNA/RNA is now ready for downstream applications. RNA extract may be stored in RNase-free water at -80°C for 1 year.

RT-ddPCR

- 7 Quantification by Droplet Digital PCR (ddPCR)

dx.doi.org/10.17504/protocols.io.bpg6mjze