MAY 14, 2019

🌐 UC Davis - Echocardiography

Anil Singapuri¹

¹University of California, Davis

Mouse Metabolic Phenotyping Centers

Tech. support email: info@mmpc.org

Lili Liang

OPEN ACCESS

DOI:

dx.doi.org/10.17504/protocols.io.yiafuae

External link:

<https://mmpc.org/shared/document.aspx?id=115&docType=Protocol>

Protocol Citation: Anil Singapuri 2019. UC Davis - Echocardiography.

protocols.io

<https://dx.doi.org/10.17504/protocols.io.yiafuae>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Feb 25, 2019

Last Modified: May 14, 2019

PROTOCOL integer ID:
20770

ABSTRACT

Summary:

Cardiac hypertrophy is one of the most common causes of heart failure. The development of cardiac hypertrophy and failure can be monitored using noninvasive imaging with echocardiogram in conscious animals.

Modified from: Timofeyev V. et al. J Mol Cell Cardiol. 2006 Jul;41(1):170-81.

Keywords: Cardiac hypertrophy, heart failure, Echocardiography

- 1 The animals' chest will be shaved and then be placed in a supine position on a warmed (37°C) towel covered surgical table.
- 2 The mice are restrained in a 50 ml conical tube with an opening for the nose and chest, a standard procedure for imaging mice for this kind of study.
- 3 The imaging generally lasts for only 2-5 min, we were trained in this method by CLAS personnel and in our experience works very well.
- 4 M-mode and two-dimensional measurements are used to assess systolic function.
- 5 An average of six selected cardiac cycles from at least two separate scans performed in randomblindfashion with papillary muscles used as a point of reference.
- 6 End diastole was defined as the maximal left ventricular (LV) diastolic dimension and end systole was defined as the peak of posterior wall motion.
- 7 Fractional shortening (FS), a surrogate of systolic function, was calculated from LV dimensions as follows:
$$FS = ((EDD - ESD) / EDD) \times 100\%$$
where EDD and ESD are LV end diastolic and end systolic dimension, respectively.