MAY 14, 2019

UC Davis - Glucose Protocol

Peter Havel¹

¹University of California, Davis

Mouse Metabolic Phenotyping Centers
Tech. support email: info@mmpc.org

Lili Liang

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.yj9fur6

External link:
<https://mmpc.org/shared/document.aspx?id=88&docType=Protocol>

Protocol Citation: Peter Havel 2019. UC Davis - Glucose Protocol.
protocols.io
<https://dx.doi.org/10.17504/protocols.io.yj9fur6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Feb 26, 2019

Last Modified: May 14, 2019

PROTOCOL integer ID:
20833

ABSTRACT

Summary:

Glucose is oxidized by glucose oxidase to gluconic acid and hydrogen peroxide. The hydrogen peroxide reacts in the presence of peroxidase with HBA and 4-aminoantipyrine forming a red quinoneimine dye. The intensity of the color formed is proportional to the glucose concentration and can be measured photometrically between 460 and 560 nm.

MATERIALS

MATERIALS

☒ Calibrator **Fisher Diagnostics Catalog #TR1591-030**

☒ Reagents **Fisher Diagnostics Catalog #TR15103**

☒ PBS **Contributed by users**

☒ Microplate **Contributed by users**

☒ Platereader **Contributed by users**

Reagent Preparation:

PBS – ready to use

Reagent – reconstitute with distilled water to make a 2X solution

Keywords: glucose

1 Reconstitute powdered reagent with only 25 ml of distilled water to make a 2X solution.

2 Add 3 µl of calibrator and sample to each well.

3 Add 150 µl of PBS to each well. Read at 540 nm.

IMPORTANT: Make sure not to add any bubbles to the wells when dispensing reagents, this will interfere with reading in the platereader.

4 Add 150 µl of 2X reagent to each well. Incubate at 37°C for 10 minutes. Read at 540 nm.

5 Subtract blank readings from final readings. The assay will be linear so the unknown samples can be calculated as $(\text{Sample Absorbance} \div \text{Calibrator Absorbance}) \times \text{Calibrator Concentration}$.