

May 15, 2019

UC Davis - Lipoprotein profiling and partical size

DOI

dx.doi.org/10.17504/protocols.io.yrifv4e



Peter Havel¹

¹University of California, Davis

Mouse Metabolic Phenotyping Centers
Tech. support email: info@mmpc.org



Lili Liang

Create & collaborate more with a free account

Edit and publish protocols, collaborate in communities, share insights through comments, and track progress with run records.

Create free account

OPEN  ACCESS



DOI: <https://dx.doi.org/10.17504/protocols.io.yrifv4e>

External link: <https://mmpc.org/shared/document.aspx?id=263&docType=Protocol>

Protocol Citation: Peter Havel 2019. UC Davis - Lipoprotein profiling and partical size. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.yrifv4e>



License: This is an open access protocol distributed under the terms of the **Creative Commons Attribution License**, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working

We use this protocol and it's working

Created: March 01, 2019

Last Modified: May 15, 2019

Protocol Integer ID: 21002

Keywords: Lipoprotein profiling, Lipoprotein size , lipoprotein profiling, lipoprotein size, additional subclasses of lipoprotein, lipoprotein, little as10µl for full analytical testing, full analytical testing, ldl, partical size summary, little as10µl, uc davi

Abstract

Summary:

Lipoprotein profiling and Lipoprotein size will be provided for CM, VLDL, LDL and HDL as well as 20 additional subclasses of lipoprotein. Testing requires as little as10µL for full analytical testing.

Troubleshooting

Before start

Lipoprotein profiling and size determined by LipoSEARCH <http://www.lipo-search.com/eng/index.php>



- 1 The sample is loaded into a gel permeation column specifically designed for separating lipoprotein components. Then, lipoprotein is eluted from the column in order of larger particles to smaller ones.
- 2 Lipoprotein fractionated by particle size is fed into reaction coils, and the degraded products after reacting with reagents are sent to detectors. The levels of cholesterol and triglyceride are output in the form of chromatogram.
- 3 The data obtained is processed with our patented analyzing program. The final output is a composite chromatogram and numeric data on 4 major classes (CM, VLDL, LDL and HDL) and 20 subclasses.