VERSION 2

SEP 03, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.63hhgj6

External link:
<https://mmpc.org/shared/document.aspx?id=123&docType=Protocol>

Protocol Citation: Peter Havel 2019. UC Davis - Protein Carbonyl. **protocols.io** <https://dx.doi.org/10.17504/protocols.io.63hhgj6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Sep 03, 2019

Last Modified: Sep 03, 2019

PROTOCOL integer ID:
 27465

UC Davis - Protein Carbonyl V.2

Peter Havel¹

¹University of California, Davis

Mouse Metabolic Phenotyping Centers
 Tech. support email: info@mmpc.org

Lili Liang

ABSTRACT

Summary:

Cayman Chemical's Protein Carbonyl Colorimetric Assay Kit is a convenient colorimetric assay for the measurement of oxidized proteins. Protein samples are derivatized by making use of the reaction between 2,4-dinitrophenylhydrazine (DNPH) and protein carbonyls. Formation of a Schiff base produces the corresponding hydrazone which can be analyzed spectrophotometrically at 360-385 nm. This assay can be used to measure oxidized protein in plasma, serum, cell lysates, and tissue homogenates.

MATERIALS

MATERIALS

⊗ HCl Contributed by users

⊗ DNPH Contributed by users

⊗ TCA Contributed by users

⊗ EtOH Contributed by users

⊗ Ethyl Acetate Contributed by users

⊗ Assay Kit Cayman Chemical Company Catalog #10005020

Note:

Cayman Chemical [RRID:SCR_008945](#)

Keywords: Protein Carbonyl

- 1 Transfer 200 µl of sample to two 2 ml plastic tubes. One tube will be the sample tube (S#) and the other will be the control tube (C#).
- 2 Add 800 µl of DNPH to the sample tube and add 800 µl of 2.5 M HCl to the control tube.
- 3 Incubate both tubes (S# & C#) in the dark at room temperature for one hour. Vortex each tube briefly every 15 minutes during the incubation.
- 4 Add 1 ml of 20% TCA to each tube and vortex. Place tubes on ice and incubate for five minutes.
- 5 Centrifuge tubes at 10,000 xg for 10 minutes at 4°C in a microcentrifuge.
- 6 Discard the supernatant and resuspend the pellet in 1 ml of 10% TCA. Place tubes on ice and let sit for five minutes.
- 7 Centrifuge tubes at 10,000 xg for 10 minutes at 4°C in a microcentrifuge.
- 8 Discard the supernatant and resuspend the pellet in 1 ml of (1:1) Ethanol/Ethyl Acetate mixture. Manually suspend pellet with spatula, vortex thoroughly, and centrifuge tubes at 10,000 xg for 10 minutes at 4°C in a microcentrifuge.

- 9 Repeat Step 8 two more times.
- 10 After the final wash, resuspend the protein pellets in 500µl of guanidine hydrochloride by vortexing.
- 11 Centrifuge tubes at 10,000 xg for 10 minutes at 4°C in a microcentrifuge to remove any left over debris.
- 12 Transfer 220 µl of supernatant from the sample (S#) tube to two wells of the 96-well plate.
- 13 Transfer 220 µl of supernatant from the control (C#) tube to two wells of the 96-well plate.
- 14 Measure the absorbance at a wavelength between 360-385 nm using a plate reader.
- 15 **Calculation**
 1. Calculate the average absorbance of each sample and control.
 2. Subtract the average absorbance of the controls from the average absorbance of the samples. This is the Corrected Absorbance (CA).
 3. Determine the concentration of the carbonyls by inserting the corrected absorbance into the following equation:

$$\text{Protein Carbonyl (nmol/ml)} = [(CA)/(*0.011 \mu\text{M}^{-1})](500 \mu\text{l}/200 \mu\text{l})$$

*The actual extinction coefficient for dinitrophenylhydrazine at 370 nm is $22,000 \text{ M}^{-1}\text{cm}^{-1}$ ($0.022 \mu\text{M}^{-1}\text{cm}^{-1}$). This value has been adjusted for the pathlength of the solution in the well.