MAY 16, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.yyifxue

External link:
<https://mmpc.org/shared/document.aspx?id=225&docType=Protocol>

Protocol Citation: Chee Lim 2019. Vandy - Mouse Echocardiography.
protocols.io
<https://dx.doi.org/10.17504/protocols.io.yyifxue>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Mar 08, 2019

Last Modified: May 16, 2019

PROTOCOL integer ID:
 21226

Vandy - Mouse Echocardiography

Chee Lim¹

¹Vanderbilt University

Mouse Metabolic Phenotyping Centers
 Tech. support email: info@mmpc.org

Lili Liang

ABSTRACT

Summary:

Transthoracic mouse echocardiography is used to provide noninvasive imaging of the heart and allows for quantification of myocardial wall and chamber dimensions and systolic and diastolic performance.

MATERIALS

MATERIALS

- Isoflurane **Contributed by users**
- Coupling gel **Contributed by users**
- Vevo2100 Imaging System **VisualSonics Catalog #Vevo2100**

- 1 The chest hair is first shaved and a topical delipatory agent (e.g. Nair) is used to remove any remaining body hair.
- 2 The conscious mouse is held in the prone position, decreasing vagal reflexes and associated abnormalities of heartrate or AV conduction.
If anesthesia is required, 3% isoflurane is used for sedation and the mouse is placed in a supine position on a heated platform (to maintain body temperature) with embedded ECG leads (Visual Sonics). A nose cone is placed over the snout and isoflurane (1%) is delivered to maintain sedation throughout the procedure.
- 3 Ultrasound coupling gel heated to 34C is applied to the chest area and a linear array transducer (18-23 MHz) is positioned to obtain two-dimensional B-mode parasternal long and short axis views at the mid-ventricular level(Vevo 2100, VisualSonics).
- 4 One-dimensional M-mode images are obtained for measurement in the short axis view to obtain cardiac wall and chamber dimensions.
- 5 Digital images are permanently archived. Depending on the type of measurement, the imaging procedure can last from one to several minutes.
- 6 Left ventricular (LV) chamber size and wall thickness are measured off-line in the M-mode from at least three consecutive beats and averaged. LV wall thickness: intraventricular septum (IVS) and posterior wall (PW) at systole and diastole; and LV internal dimensions (LVID) during systole and diastole are measured. LV percent fractional shortening (FS) and ejection fraction (EF) are calculated from the M-mode measurements.