MAY 16, 2019

Yale - Beta hydroxybutyrate (Cobas)

Gary Cline¹, John Stack¹

¹Yale University

Mouse Metabolic Phenotyping Centers
Tech. support email: info@mmpc.org

Lili Liang

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.y2bfyan

External link:
<https://mmpc.org/shared/document.aspx?id=215&docType=Protocol>

Protocol Citation: Gary Cline, John Stack 2019. Yale - Beta hydroxybutyrate (Cobas). **protocols.io** <https://dx.doi.org/10.17504/protocols.io.y2bfyan>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Mar 11, 2019

Last Modified: May 16, 2019

PROTOCOL integer ID:
21283

ABSTRACT

Summary:

Procedure used to determine the concentration of β -Hydroxybutyrate in blood, serum, and plasma. β -Hydroxybutyrate is measured by the oxidation to acetoacetate, followed by reduction of an indicator dye (monitored at 505 nm) by NADH.

MATERIALS

MATERIALS

β -Hydroxybutyrate Controls **Pointe Scientific Inc. Catalog #H7587-CTL**

β -Hydroxybutyrate Reagent (test kit includes standard). **Pointe Scientific Inc. Catalog #H7587-58**

Reagent Preparation:

β -Hydroxybutyrate Controls: As supplied by vendor

β -Hydroxybutyrate Reagent and Standard: As supplied by vendor

BEFORE START INSTRUCTIONS

Analysis by automated system Cobas Mira Plus.

Keywords: Beta
hydroxybutyrate

- 1 Calibrate Cobas for β -Hydroxybutyrate by running a β -Hydroxybutyrate standard and three β -Hydroxybutyrate controls.

- 2 Sample handling as performed by the Cobas Mira Plus.
 - a) Pipette 3 μ L of sample into cuvette.
 - b) Add 105 μ L of β -Hydroxybutyrate Reagent.
 - c) Mixture is incubated at 37°C for 10 minutes.
 - d) Absorbance is measured at 505 nm.