MAY 17, 2019

Yale - LDL Cholesterol

John Stack¹, Gary Cline¹

¹Yale University

Mouse Metabolic Phenotyping Centers
Tech. support email: info@mmpc.org

Lili Liang

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.y3vfyn6

External link:
<https://mmpc.org/shared/document.aspx?id=211&docType=Protocol>

Protocol Citation: John Stack, Gary Cline 2019. Yale - LDL Cholesterol.
protocols.io
<https://dx.doi.org/10.17504/protocols.io.y3vfyn6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
We use this protocol and it's working

Created: Mar 12, 2019

Last Modified: May 17, 2019

PROTOCOL integer ID:
21333

ABSTRACT

Summary:

Procedure used to determine the concentration of LDL cholesterol in blood, serum, and plasma. LDL-cholesterol is determined by an elimination method to selectively remove HDL, VLDL, and IDL subfractions prior to assay of LDL-cholesterol as described for total cholesterol.

MATERIALS

MATERIALS

Lipid Calibrator Prolabs(cliniqa) Catalog #R85528

LDL Direct Reagent 1 Prolabs(cliniqa) Catalog #R85556

LDL Direct Reagent 2 Prolabs(cliniqa) Catalog #R85556

Reagent Preparation:

LDL Direct Reagent 1: As supplied by vendor.

LDL Direct Reagent 2: As supplied by vendor.

Lipid Calibrator: Add 1mL of water to lipid calibrator bottle. Invert to mix, allowing 15 minutes for the reagent to settle. For the calibration, prepare serial dilutions.

BEFORE START INSTRUCTIONS

Analysis by automated system Cobas Mira Plus

Keywords: LDL Cholesterol

- 1** Calibrate Cobas for LDL by running dilutions of the lipid calibrator and the addition of the two direct reagents.

- 2** Sample handling as performed by the Cobas Mira Plus.
 - a) Pipette 3 μ L of sample into a cuvette slot.
 - b) Add 225 μ L of LDL Direct Reagent 1.
 - c) Add 75 μ L of LDL Direct Reagent 2.
 - d) Mixture is incubated at 37°C and spun for 10 minutes.
 - e) Absorbance is measured at 600 nm.