MAY 17, 2019

OPEN ACCESS

DOI:
dx.doi.org/10.17504/protocols.io.y3zfyp6

External link:
<https://mmpc.org/shared/document.aspx?id=218&docType=Protocol>

Protocol Citation: John Stack, Gary Cline 2019. Yale - Magnesium. **protocols.io**
<https://dx.doi.org/10.17504/protocols.io.y3zfyp6>

License: This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](#), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited

Protocol status: Working
 We use this protocol and it's working

Created: Mar 12, 2019

Last Modified: May 17, 2019

PROTOCOL integer ID:
 21337

Keywords: Magnesium

Yale - Magnesium

John Stack¹, Gary Cline¹

¹Yale University

Mouse Metabolic Phenotyping Centers
 Tech. support email: info@mmpc.org

Lili Liang

ABSTRACT

Summary:

Procedure used to determine the concentration of magnesium in blood, serum, and plasma. Magnesium is measured as the chlorophosphonazo III complex (addition of EGTA prevents calcium interference) and monitored at 600 nm.

MATERIALS

MATERIALS

⊗ Magnesium Reagent **Prolabs(cliniqa) Catalog #125-12**

⊗ Magnesium Calibrator **Prolabs(cliniqa) Catalog #125-12**

Reagent Preparation:

Magnesium Reagent: As supplied by vendor

Magnesium Calibrator: As supplied by vendor

BEFORE START INSTRUCTIONS

Analysis by automated system Cobas Mira Plus

- 1 Calibrate Cobas for Magnesium analysis by running the Magnesium Calibrator and a Water Blank.

- 2 Sample handling as performed by the Cobas Mira Plus
 - a) Pipette 4 μL of sample into a cuvette slot.
 - b) Add 350 μL of Magnesium Reagent.
 - c) Mixture is incubated at 37°C for 10 minutes
 - d) Absorbance is measured at 600 nm.